# $Hjerne\_PiB\_PET\_LM\_5.MlAdult\_PET5$

30. maj 2013

## Indhold

1	T O			
	1.1	Routine	1	
	1.2	Scan	1	
<b>2</b>	$\mathbf{CT}$	Brain	1	
	2.1	Routine	1	
	2.2	Scan	1	
	2.3	Recons	2	
		2.3.1 Recon 1	2	
3	Pau	se	2	
4	PET	Γ Brain LM	2	
	4.1	Routine	2	
	4.2	Scan	2	
	4.3	Recons	3	
		4.3.1 Recon 1	3	

## 1 Topogram

#### 1.1 Routine

mA: 35kV: 120

Topogram length: 256 mmTube position: Lateral

#### 1.2 Scan

mA: 35kV: 120Delay: 4s

Topogram length: 256 mmDirection: CraniocaudalTube position: Lateral

• API: None

• Kernel: 20s standar

• Window: Topogram Head

## 2 CT Brain

#### 2.1 Routine

• Eff. mAs: 40

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

CTDlvol: 5.504mGy
Scan time: 9.310 s
Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

#### 2.2 Scan

• Quality ref. mAs: 380

• Eff. mAs: 40

• kV: 120

Scan time: 9.310 sRotation time: 1.000 s

Delay: 4.000 sSlice: 3.00 mmPitch: 1.0000000

• Direction: Craniocaudal

#### 2.3 Recons

#### 2.3.1 Recon 1

- Series description: AC CT
- Slice: 3.00
- Kernel: H19s PET very smooth
- Window: CerebrumExtended FoV: Off
- FoV: 300
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 74

## 3 Pause

## 4 PET Brain LM

#### 4.1 Routine

- Isotope: C-11
- Pharm.: PIB
- Inj. Dose: 700 MegaBequerels
- Scan mode: List mode
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 30 Minutes

#### 4.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: List mode
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

### 4.3 Recons

#### 4.3.1 Recon 1

• Series description: PET Brain LM

Recon range (bed): -1 to -1Output image type: NoRecon

• Recon method: Iterative

Iterations: 6 Subsets: 16 Image size: 256

• Zoom: 2

Filter: GaussianFWHM (mm): 2Offset X: 0 mmOffset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: On

Match CT slice location: On Save intermediate data: Off