$4_1_PET_CT_WB_FDG_uden_IV.Adult$

22. oktober 2012

Indhold

1	Top	Topogram														1							
	1.1	Routin	e														 						1
	1.2	Scan															 						1
2	\mathbf{CT}	WB																					1
	2.1	Routin	ie														 						1
	2.2	Scan															 						2
	2.3	Recons	3														 						2
		2.3.1	Recon 1																				2
		2.3.2	Recon 2																				3
		2.3.3	Recon 3																				4
		2.3.4																					4
		2.3.5	Recon 5																				5
3	Pau	se																					6
4	PET WB														6								
	4.1	Routin	ie														 						6
	4.2	Scan															 						6
	4.3		5																				7
		4.3.1	Recon 1																				7
		4.3.2	Recon 2																				7

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 1024 mm
- Tube position: Top

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 1024 mm
- Direction: Craniocaudal
- Tube position: Top
- API: API
- Kernel: 80
- Window: Topogram

2 CT WB

2.1 Routine

- Eff. mAs: 225
- kV: 120
- CARE Dose4D: On
- CareDoseType: CareDoseAEC
- CTDlvol: 16.38mGy

• Scan time: 32.530 s

• Delay: 2.000 s

• Slice: 2 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

2.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 225

• kV: 120

 \bullet Scan time: 32.530 s

• Rotation time: 0.500 s

 \bullet Delay: 2.000 s

• Slice: 2 mm

• Pitch: 0.80

• Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

• Series description: tmp CT WB RTD

• Slice: 4

• Kernel: B19f

• Window: Abdomen

• Extended FoV: false

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 4.000

• No. of images: 245

2.3.2 Recon 2

• Series description: AC CT 2 mm

• Slice: 2

• Kernel: B19f

• Window: Abdomen

• Extended FoV: true

• FoV: 780

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 2.000

• No. of images: 489

2.3.3 Recon 3

• Series description: CT WB 2.0 B30f

• Slice: 2

• Kernel: B30f

• Window: Abdomen

• Extended FoV: false

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 2.000

• No. of images: 489

2.3.4 Recon 4

• Series description: CT WB 2.0 cor B30f

• Slice: 2

• Kernel: B30f

• Window: Abdomen

• Extended FoV: false

• FoV: 977

• Center X: 0

• Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: 3D
- Recon Axis: Coronal
- Image order: AP
- Recon increment: 2.000
- No. of images: 251

2.3.5 Recon 5

- Series description: CT WB 2.0 sag B30f
- Slice: 2
- Kernel: B30f
- Window: Abdomen
- Extended FoV: false
- FoV: 977
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: 3D
- Recon Axis: Sagittal
- ullet Image order: LeftToRight
- Recon increment: 2.000
- No. of images: 489

3 Pause

4 PET WB

4.1 Routine

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- Inj. Dose: 283000000.000000 Bequerels
- Inj. date (date/month year): 14/05 2012
- Inj. time: 10:39:23
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 2.000000 Minutes

4.2 Scan

- Autoload: true
- Rebinner LUT: N/A
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- ULD (keV): N/A

4.3 Recons

4.3.1 Recon 1

- Series description: PET WB Image check HD
- Recon range (bed): 1 to 7
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 2
- Subsets: 12
- Image size: 200
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false

4.3.2 Recon 2

- Series description: PET WB Corr TrueX+TOF 2 mm
- Recon range (bed): 1 to 7
- Output image type: Corrected
- Recon method: TrueXTOF
- Iterations: 2
- Subsets: 21

• Image size: 400

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 2.000000

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: true

• Match CT slice location: true

• Save intermediate data: false