

4_4_PET_CT_FDG_Brain.Adult

22. oktober 2012

Indhold

1	Topogram	1
1.1	Routine	1
1.2	Scan	1
2	CT Brain	1
2.1	Routine	1
2.2	Scan	2
2.3	Recons	2
2.3.1	Recon 1	2
2.3.2	Recon 2	3
2.3.3	Recon 3	4
3	Pause	4
4	PET Brain	4
4.1	Routine	4
4.2	Scan	5
4.3	Recons	5
4.3.1	Recon 1	5
4.3.2	Recon 2	6

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 256 mm
- Tube position: Lateral

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 256 mm
- Direction: Craniocaudal
- Tube position: Lateral
- API: API
- Kernel: 20
- Window: Topogram

2 CT Brain

2.1 Routine

- Eff. mAs: 380
- kV: 120
- CARE Dose4D: On
- CareDoseType: CareDoseAEC
- CTDIvol: 64.904mGy

- Scan time: 69.330 s
- Delay: 2.000 s
- Slice: 2 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

2.2 Scan

- Quality ref. mAs: 380
- Eff. mAs: 380
- kV: 120
- Scan time: 69.330 s
- Rotation time: 1.000 s
- Delay: 2.000 s
- Slice: 2 mm
- Pitch: 0.55
- Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

- Series description: tmp CT Brain RTD
- Slice: 5
- Kernel: H19s
- Window: Cerebrum
- Extended FoV: false
- FoV: 300
- Center X: 0

- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 5.000
- No. of images: 45

2.3.2 Recon 2

- Series description: AC CT
- Slice: 2
- Kernel: H19s
- Window: Cerebrum
- Extended FoV: false
- FoV: 300
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 112

2.3.3 Recon 3

- Series description: Cranie VIR 2.0 H30S
- Slice: 2
- Kernel: H30s
- Window: Cerebrum
- Extended FoV: false
- FoV: 300
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 112

3 Pause

4 PET Brain

4.1 Routine

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- Inj. Dose: 200000000.000000 Bequerels
- Inj. date (date/month - year): 08/08 - 2012
- Inj. time: 14:22:13

- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 10.000000 Minutes

4.2 Scan

- Autoload: true
- Rebinner LUT: N/A
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- ULD (keV): N/A

4.3 Recons

4.3.1 Recon 1

- Series description: PET Brain Corrected 3 mm zoom 3
- Recon range (bed): 1 to 1
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 24
- Image size: 400
- Zoom: 3.000000
- Filter: Gaussian
- FWHM (mm): 3.000000

- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false

4.3.2 Recon 2

- Series description: Neurostat
- Recon range (bed): 1 to 1
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 8
- Image size: 128
- Zoom: 3.000000
- Filter: Gaussian
- FWHM (mm): 8.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false