4_PETCT_AC_Brain_Ped.Child

8. oktober 2012

Indhold

1 Topogram

1.1 Routine

- mA: 50

- kV: 80

- Topogram length: 256 mm

- Tube position: Lateral

1.2 Scan

- mA: 50

- kV: 80

- Delay: 4s

- Topogram length: 256 mm

- Direction: Craniocaudal

- Tube position: Lateral

- API:

- Kernel: T20s

- Window: Topogra

2 CT WB

2.1 Routine

- Eff. mAs: 85

- kV: 80

- CARE Dose4D: Off

- CTDlvol: 4.4795mGy

- Scan time: 20.050 s

- Delay: 2.000 s

- Slice: 3 mm

- No. of images: Samme som i foerste recon, slet?(y/n)

- Tilt: 0.0 grader

2.2 Scan

- Quality ref. mAs: 85

- Eff. mAs: 85

- kV: 80

- Scan time: 20.050 s

- Rotation time: 1.000 s

- Delay: 2.000 s

- Slice: 3 mm

- Pitch: 1.00

- Direction: Craniocaudal

2.3 Recon

2.3.1 Recon 1

- Series description: tmp CT Brain RTD

- Slice: 4

- Kernel: "H19s"

- Window: "Cerebrum"

- Extended FoV: false

- FoV: 300

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Caudocranial

- Recon increment: 4.000

- No. of images: 56

2.3.2   Recon 2

- Series description: AC CT Brain

- Slice: 3

- Kernel: "H19s"

- Window: "Cerebrum"

- Extended FoV: false

- FoV: 300

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Caudocranial

- Recon increment: 1.500

- No. of images: 148

3 Pause

4 PET WB

4.1 Routine

- Isotope: F-18

- Pharm.: FDG

- Inj. Dose: 9

- Inj. date (date/month - year): 06/10 - 2008

- Inj. time: 12:18:30

- Scan mode: Sinogram

- Scan range: Match CT FOV

- No. of beds: N/A

- Scan duration/bed: 10.000000 Minutes

4.2 Scan

- Autoload: true

- Rebinner LUT: N/A

- Scan output: Sinogram

- Sinogram mode: PromptsRandoms

- Input trigger signal: None

- LLD (keV): N/A

- ULD (keV): N/A

4.3   Recon

4.3.1   Recon 1

- Series description: PET Brain

- Recon range (bed): ups..

- Output image type: Corrected

- Recon method: Iterative3D

- Iterations: 4

- Subsets: 24

- Image size: 400

- Zoom: 2.000000

- Filter: Gaussian

- FWHM (mm): 4.000000

- Offset X: 0 mm

- Offset Y: 0 mm

- Attenuation correction: On (1)

- Scatter correction: true

- Match CT slice location: true

- Save intermediate data: false