

# 4\_PET\_CT\_LM\_FET\_Brain\_Terapi\_DYN.Adult

23. oktober 2012

# Indhold

|          |                     |          |
|----------|---------------------|----------|
| <b>1</b> | <b>Topogram</b>     | <b>1</b> |
| 1.1      | Routine . . . . .   | 1        |
| 1.2      | Scan . . . . .      | 1        |
| <b>2</b> | <b>CT Brain</b>     | <b>1</b> |
| 2.1      | Routine . . . . .   | 1        |
| 2.2      | Scan . . . . .      | 2        |
| 2.3      | Recons . . . . .    | 2        |
| 2.3.1    | Recon 1 . . . . .   | 2        |
| 2.3.2    | Recon 2 . . . . .   | 3        |
| 2.3.3    | Recon 3 . . . . .   | 4        |
| 2.3.4    | Recon 4 . . . . .   | 4        |
| <b>3</b> | <b>Pause</b>        | <b>5</b> |
| <b>4</b> | <b>PET Brain LM</b> | <b>5</b> |
| 4.1      | Routine . . . . .   | 5        |
| 4.2      | Scan . . . . .      | 5        |
| 4.3      | Recons . . . . .    | 6        |
| 4.3.1    | Recon 1 . . . . .   | 6        |
| 4.3.2    | Recon 2 . . . . .   | 7        |
| 4.3.3    | Recon 3 . . . . .   | 7        |
| 4.3.4    | Recon 4 . . . . .   | 8        |
| 4.3.5    | Recon 5 . . . . .   | 9        |

# **1 Topogram**

## **1.1 Routine**

- mA: 35
- kV: 120
- Topogram length: 256 mm
- Tube position: Lateral

## **1.2 Scan**

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 256 mm
- Direction: Craniocaudal
- Tube position: Lateral
- API: API
- Kernel: 20
- Window: Topogram

# **2 CT Brain**

## **2.1 Routine**

- Eff. mAs: 340
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDIvol: 58.072mGy

- Scan time: 48.120 s
- Delay: 2.000 s
- Slice: 1 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

## **2.2 Scan**

- Quality ref. mAs: 350
- Eff. mAs: 340
- kV: 120
- Scan time: 48.120 s
- Rotation time: 1.000 s
- Delay: 2.000 s
- Slice: 1 mm
- Pitch: 0.80
- Direction: Caudocranial

## **2.3 Recons**

### **2.3.1 Recon 1**

- Series description: tmp CT Brain RTD
- Slice: 5
- Kernel: H19s
- Window: Cerebrum
- Extended FoV: false
- FoV: 300
- Center X: 0

- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 5.000
- No. of images: 45

### **2.3.2 Recon 2**

- Series description: AC CT 1mm Brain
- Slice: 1
- Kernel: H19s
- Window: Cerebrum
- Extended FoV: false
- FoV: 300
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 1.000
- No. of images: 224

### **2.3.3 Recon 3**

- Series description: CT 3.0mm Brain H31s
- Slice: 3
- Kernel: H31s
- Window: Cerebrum
- Extended FoV: false
- FoV: 250
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 3.000
- No. of images: 75

### **2.3.4 Recon 4**

- Series description: AC CT8mm Brain
- Slice: 8
- Kernel: H31s
- Window: Cerebrum
- Extended FoV: false
- FoV: 300
- Center X: 0
- Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 8.000
- No. of images: 28

### **3    Pause**

## **4    PET Brain LM**

### **4.1   Routine**

- Isotope: F-18
- Pharm.: FET
- Inj. Dose: 205000000.000000 Bequerels
- Inj. date (date/month - year): 19/09 - 2012
- Inj. time: 12:52:55
- Scan mode: List mode
- Scan range: Match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 40.000000 Minutes

### **4.2   Scan**

- Autoload: false
- Rebinner LUT: N/A
- Scan output: List mode

- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- ULD (keV): N/A

## **4.3 Recons**

### **4.3.1 Recon 1**

- Series description: FET dyn 20-40 5mm CT3mm
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400
- Zoom: 2.500000
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (2)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false



### 4.3.2 Recon 2

- Series description: FET dyn 10-30min 5mm CT8mm
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400
- Zoom: 2.500000
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (3)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false

### 4.3.3 Recon 3

- Series description: FET 8Fdyn 5 mm CT3mm 4
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400

- Zoom: 2.500000
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (2)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false

#### **4.3.4 Recon 4**

- Series description: FET 22Fdyn 5 mm CT3mm 4
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400
- Zoom: 2.500000
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (2)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false

#### 4.3.5 Recon 5

- Series description: FET dyn 25-35 5mm CT3mm
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400
- Zoom: 2.500000
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (2)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false