

TUMOR_HUND_ATSM.MlAdult_PET5

7. marts 2013

Indhold

| | | |
|----------|-------------------|----------|
| 1 | Topogram | 1 |
| 1.1 | Routine | 1 |
| 1.2 | Scan | 1 |
| 2 | CT TUMOR | 1 |
| 2.1 | Routine | 1 |
| 2.2 | Scan | 1 |
| 2.3 | Recons | 2 |
| 2.3.1 | Recon 1 | 2 |
| 2.3.2 | Recon 2 | 2 |
| 3 | Pause | 2 |
| 4 | PET STATIC | 2 |
| 4.1 | Routine | 2 |
| 4.2 | Scan | 3 |
| 4.3 | Recons | 3 |
| 4.3.1 | Recon 1 | 3 |
| 4.3.2 | Recon 2 | 3 |

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 0 mm
- Tube position: Top

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 0 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: 80s shar
- Window: Topogram Body

2 CT TUMOR

2.1 Routine

- Eff. mAs: 170
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDIvol: 11.4448mGy
- Scan time: 4.000 s
- Delay: 4.000 s
- Slice: 3.00 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

2.2 Scan

- Quality ref. mAs: 170
- Eff. mAs: 170
- kV: 120
- Scan time: 4.000 s
- Rotation time: 0.500 s
- Delay: 4.000 s
- Slice: 3.00 mm
- Pitch: 1.2000000

- Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

- Series description: AC CT
- Slice: 3.00
- Kernel: B19f PET very smooth
- Window: Abdomen
- Extended FoV: On
- FoV: 700
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 74

2.3.2 Recon 2

- Series description: CT TUMOR 3.0 B30f
- Slice: 3.00
- Kernel: B30f medium smooth
- Window: Abdomen
- Extended FoV: Off
- FoV: 500
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 74

3 Pause

4 PET STATIC

4.1 Routine

- Isotope: Cu-64

- Pharm.: ATSM
- Inj. Dose: 226 MegaBequerels
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 20 Minutes

4.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

4.3 Recons

4.3.1 Recon 1

- Series description: PET TUMOR Image check
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 3
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

4.3.2 Recon 2

- Series description: PET TUMOR True X
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 3
- Subsets: 21

- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: Off
- Match CT slice location: On
- Save intermediate data: Off