# $4\_PET\_CT\_WB\_FDG\_terapi\_uden\_IV.Adult\_PET4$

20. februar 2013

## Indhold

1	$\operatorname{Top}$	Topogram														1															
	1.1	Routin	e																												1
	1.2	Scan .																													1
2	$\mathbf{CT}$	WB																													1
	2.1	Routin	e																												1
	2.2	Scan .																													2
	2.3	Recons	3																												2
		2.3.1	Recon 1																												2
		2.3.2	Recon 2																												3
		2.3.3																													4
		2.3.4	Recon 4																												4
		2.3.5	Recon 5																												5
3	Pau	se																													6
4	4 PET WB																6														
	4.1	Routin	e																												6
	4.2	Scan .																													6
	4.3	Recons																													6
	2.0	4.3.1	Recon 1																												6
		-	Recon 2	• •	•	•	•		•	•	• •	•	•	•	•	•	•	•	•	•	•	•	•	•	•	•	•	•	•	•	7

## 1 Topogram

## 1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 1024 mm
- Tube position: Top

#### 1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 1024 mm
- Direction: Craniocaudal
- Tube position: Top
- API: API
- Kernel: 80
- Window: Topogram

## 2 CT WB

## 2.1 Routine

- Eff. mAs: 225
- kV: 120
- CARE Dose4D: On
- CareDoseType: CareDoseAEC
- CTDlvol: m 1=0.000000 16.4451mGy

• Scan time: 32.745 s

• Delay: 2.000 s

• Slice: 2 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

## 2.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 225

• kV: 120

• Scan time: 32.745 s

• Rotation time: 0.500 s

• Delay: 2.000 s

• Slice: 2 mm

• Pitch: m 1=-1 0.80

• Direction: Craniocaudal

## 2.3 Recons

#### 2.3.1 Recon 1

 $\bullet$  Series description: m 1=12345MA m 2= CT WB RTD

• Slice: 4

• Kernel: B19f

• Window: Abdomen

• Extended FoV: Off

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 4.000

• No. of images: 245

#### 2.3.2 Recon 2

• Series description: m 1=45AC CT 2 mm MA m 2= AC CT 2 mm

• Slice: 2

• Kernel: B19f

• Window: Abdomen

• Extended FoV: On

• FoV: 780

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 2.000

• No. of images: 489

#### 2.3.3 Recon 3

- $\bullet$  Series description: m 1=12345MA m 2= CT WB 2.0 B30f
- Slice: 2
- Kernel: B30f
- Window: Abdomen
- Extended FoV: Off
- FoV: 500
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 489

#### 2.3.4 Recon 4

- Series description: m 1=1245B30f MA m 2= CT WB 2.0 cor B30f
- Slice: 2
- Kernel: B30f
- Window: Abdomen
- Extended FoV: Off
- FoV: 977
- Center X: 0
- Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: 3D
- Recon Axis: Coronal
- Image order: AP
- Recon increment: 2.000
- No. of images: 251

#### 2.3.5 Recon 5

- $\bullet$  Series description: m 1=1245B30f MA m 2= CT WB 2.0 sag B30f
- Slice: 2
- Kernel: B30f
- Window: Abdomen
- Extended FoV: Off
- FoV: 977
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: 3D
- Recon Axis: Sagittal
- ullet Image order: LeftToRight
- Recon increment: 2.000
- No. of images: 489

## 3 Pause

## 4 PET WB

## 4.1 Routine

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- Inj. Dose: 300000000 Bequerels
- Scan mode: Sinogram
- Scan range: Do not match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 2.000000 Minutes

## 4.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): n/a
- ULD (keV): n/a

#### 4.3 Recons

#### 4.3.1 Recon 1

- Series description: PET WB Image check HD
- Recon range (bed): 1 to 7
- Output image type: Corrected

• Recon method: TrueX

• Iterations: 2

• Subsets: 12

• Image size: 200

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 4.000000

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: Off ()

• Scatter correction: On

• Match CT slice location: On

• Save intermediate data: Off

#### 4.3.2 Recon 2

• Series description: PET WB Corr TrueX+TOF 2 mm

• Recon range (bed): 1 to 7

• Output image type: Corrected

• Recon method: TrueXTOF

• Iterations: 2

• Subsets: 21

• Image size: 400

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 2.000000

• Offset X: 0 mm

• Offset Y: 0 mm

 $\bullet$  Attenuation correction: Off ()

• Scatter correction: On

• Match CT slice location: On

• Save intermediate data: Off