# WB\_FDG\_PET\_5.MlAdult\_PET5

17. november 2012

## Indhold

1	1 Topogram	1
	1.1 Routine	 1
	1.2 Scan	 1
<b>2</b>	2 Lavdos CT WB	1
	2.1 Routine	 1
	2.2 Scan	 2
	2.3 Recons	 2
	2.3.1 Recon 1	 2
3	3 Pause	3
4	4 PET WB	3
	4.1 Routine	 3
	4.2 Scan	 4
	4.3 Recons	 4
	4.3.1 Recon 1	 4
	4.3.2 Recon 2	 5

## 1 Topogram

## 1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 1024 mm
- Tube position: Top

#### 1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 1024 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: 80s shar
- Window: Topogram Body

## 2 Lavdos CT WB

## 2.1 Routine

- Eff. mAs: 40
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDlvol: 2.70069mGy

• Scan time: 18.670 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

## 2.2 Scan

• Quality ref. mAs: 50

• Eff. mAs: 40

• kV: 120

 $\bullet$  Scan time: 18.670 s

• Rotation time: 0.500 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 0.8000000

• Direction: Craniocaudal

### 2.3 Recons

#### 2.3.1 Recon 1

• Series description: AC CT

• Slice: 3.00

• Kernel: B19f PET very smooth

• Window: Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 274

## 3 Pause

## 4 PET WB

#### 4.1 Routine

• Isotope: F-18

• Pharm.: FDG

• Inj. Dose: 400 MegaBequerels

 $\bullet\,$  Inj. date (date/month - year): 23/03 - 2011

• Inj. time: 13:06:00

• Scan mode: Sinogram

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 150 Seconds

#### 4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

• Input trigger signal: None

• LLD (keV): 435

• ULD (keV): 650

### 4.3 Recons

#### 4.3.1 Recon 1

• Series description: PET WB Corrected

• Recon range (bed): Automatic/All/dunno

• Output image type: Corrected

• Recon method: Iterative

• Iterations: 4

• Subsets: 8

• Image size: 168

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 4

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: On

• Match CT slice location: On

• Save intermediate data: Off

#### 4.3.2 Recon 2

- $\bullet\,$  Series description: PET WB Corr. True X 2 mm
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 3
- Subsets: 21
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off