$4_11_X_cal_PET4_FDG.Adult$

22. oktober 2012

Indhold

| 1 | Topogram | | | |
|---|----------------|---------------|--|--|
| | 1.1 | Routine | | |
| | 1.2 | Scan | | |
| 2 | $oldsymbol{v}$ | | | |
| | 2.1 | Routine | | |
| | 2.2 | Scan | | |
| | 2.3 | Recons | | |
| | | 2.3.1 Recon 1 | | |
| | | 2.3.2 Recon 2 | | |
| 3 | Pau | ise | | |
| 4 | PET WB | | | |
| | 4.1 | Routine | | |
| | 4.2 | Scan | | |
| | 4.3 | Recons | | |
| | | 4.3.1 Recon 1 | | |

1 Topogram

1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 256 mm

• Tube position: Top

1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 256 mm

• Direction: Craniocaudal

• Tube position: Top

• API: API

• Kernel: 80

• Window: Topogram

2 ReadOnly=1 ResID=921 AC CT Body

2.1 Routine

• Eff. mAs: 225

• kV: 120

• CARE Dose4D: On

• CareDoseType: CareDoseAEC

• CTDlvol: 16.38mGy

• Scan time: 8.050 s

• Delay: 2.000 s

• Slice: 2 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

2.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 225

• kV: 120

 \bullet Scan time: 8.050 s

• Rotation time: 0.500 s

• Delay: 2.000 s

• Slice: 2 mm

• Pitch: 0.80

• Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

• Series description: tmp AC CT Body RTD

• Slice: 4

• Kernel: B19f

• Window: Abdomen

• Extended FoV: false

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 4.000

• No. of images: 56

2.3.2 Recon 2

• Series description: AC CT 2 mm

• Slice: 2

• Kernel: B19f

• Window: Abdomen

• Extended FoV: true

• FoV: 780

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 2.000

• No. of images: 112

3 Pause

4 PET WB

4.1 Routine

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- Inj. Dose: 85000000.000000 Bequerels
- Inj. date (date/month year): 14/05 2012
- Inj. time: 10:37:14
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 5.000000 Minutes

4.2 Scan

- Autoload: true
- Rebinner LUT: N/A
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- ULD (keV): N/A

4.3 Recons

4.3.1 Recon 1

- Series description: PET WB Corr TrueX+TOF 2 mm
- Recon range (bed): 1 to 1
- Output image type: Corrected
- Recon method: TrueXTOF
- Iterations: 2
- Subsets: 21
- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false