

6_TUMOR_HYPOXI_HUND_DAG_4.MlAdult

22. oktober 2012

Indhold

1	Topogram	1
1.1	Routine	1
1.2	Scan	1
2	CT TUMOR	1
2.1	Routine	1
2.2	Scan	2
2.3	Recons	2
2.3.1	Recon 1	2
2.3.2	Recon 2	3
3	Pause	4
4	PET STATIC	4
4.1	Routine	4
4.2	Scan	4
4.3	Recons	5
4.3.1	Recon 1	5
4.3.2	Recon 2	5

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 512 mm
- Tube position: Top

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 270 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: 80s shar
- Window: Topogram Body

2 CT TUMOR

2.1 Routine

- Eff. mAs: 170
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDIvol: 11.4448mGy

- Scan time: 4.000 s
- Delay: 4.000 s
- Slice: 3.00 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 170 grader

2.2 Scan

- Quality ref. mAs: 170
- Eff. mAs: 170
- kV: 120
- Scan time: 4.000 s
- Rotation time: 0.500 s
- Delay: 4.000 s
- Slice: 3.00 mm
- Pitch: 1.2000000
- Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

- Series description: AC CT
- Slice: 3.00
- Kernel: B19f PET very smooth
- Window: Abdomen
- Extended FoV: On
- FoV: 700
- Center X: 0

- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 74

2.3.2 Recon 2

- Series description: CT TUMOR 3.0 B30f
- Slice: 3.00
- Kernel: B30f medium smooth
- Window: Abdomen
- Extended FoV: Off
- FoV: 500
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 74

3 Pause

4 PET STATIC

4.1 Routine

- Isotope: Cu-64
- Pharm.:
- Inj. Dose: 0 MegaBequerels
- Inj. date (date/month - year): 22/04 - 2008
- Inj. time: 12:03:00
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 20 Minutes

4.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

4.3 Recons

4.3.1 Recon 1

- Series description: PET TUMOR Corrected 180 min
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 256
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 3
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

4.3.2 Recon 2

- Series description: PET TUMOR Uncorrected 180 min
- Recon range (bed): Automatic/All/dunno
- Output image type: Uncorrected
- Recon method: Iterative
- Iterations: 4
- Subsets: 8

- Image size: 256
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 3
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: Off
- Match CT slice location: On
- Save intermediate data: Off