

lavdosisCTHjerne.Adult\_PET4

27. februar 2013

# Indhold

<b>1</b>	<b>Topogram</b>	<b>1</b>
1.1	Routine . . . . .	1
1.2	Scan . . . . .	1
<b>2</b>	<b>CT Brain</b>	<b>1</b>
2.1	Routine . . . . .	1
2.2	Scan . . . . .	2
2.3	Recons . . . . .	2
2.3.1	Recon 1 . . . . .	2
2.3.2	Recon 2 . . . . .	3

# **1 Topogram**

## **1.1 Routine**

- mA: 35
- kV: 120
- Topogram length: 256 mm
- Tube position: Lateral

## **1.2 Scan**

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 256 mm
- Direction: Craniocaudal
- Tube position: Lateral
- API: API
- Kernel: 20
- Window: Topogram

# **2 CT Brain**

## **2.1 Routine**

- Eff. mAs: 30
- kV: 120
- CARE Dose4D: On
- CareDoseType: CareDoseAEC
- CTDIvol: 5.12404mGy

- Scan time: 2.444 s
- Delay: 2.000 s
- Slice: 3 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

## **2.2 Scan**

- Quality ref. mAs: 30
- Eff. mAs: 30
- kV: 120
- Scan time: 2.444 s
- Rotation time: 1.000 s
- Delay: 2.000 s
- Slice: 3 mm
- Pitch: 1.50
- Direction: Craniocaudal

## **2.3 Recons**

### **2.3.1 Recon 1**

- Series description: CT Brain RTD
- Slice: 1.5
- Kernel: H19s
- Window: Cerebrum
- Extended FoV: Off
- FoV: 300
- Center X: 0

- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 1.500
- No. of images: 7

### **2.3.2 Recon 2**

- Series description: AC CT
- Slice: 3
- Kernel: H19s
- Window: Cerebrum
- Extended FoV: Off
- FoV: 300
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 1