

PETCT_Cardiac.MlAdult

1. oktober 2012

Indhold

| | | |
|----------|-------------------|----------|
| 1 | Topogram | 1 |
| 1.1 | Routine | 1 |
| 1.2 | Scan | 1 |
| 2 | CT WB | 1 |
| 2.1 | Routine | 1 |
| 2.2 | Scan | 2 |
| 2.3 | Recon | 2 |
| 2.3.1 | Recon 1 | 2 |
| 3 | Pause | 3 |
| 4 | PET WB | 3 |
| 4.1 | Routine | 3 |
| 4.2 | Scan | 3 |
| 4.3 | Recon | 4 |
| 4.3.1 | Recon 1 | 4 |

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 256 mm
- Tube position: Top

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 256 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: T80s sharp
- Window: Topogram Body

2 CT WB

2.1 Routine

- Eff. mAs: 11
- kV: 3
- CARE Dose4D: Off
- CTDIvol: 0.7414mGy
- Scan time: 1.5000000 s

- Delay: 4.000 s
- Slice: 3.00 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 11 grader

2.2 Scan

- Quality ref. mAs: 170
- Eff. mAs: 11
- kV: 120
- Scan time: 3.360 s
- Rotation time: 0.500 s
- Delay: 4.000 s
- Slice: 0.6 mm
- Pitch: 1.5000000
- Direction: Craniocaudal

2.3 Recon

2.3.1 Recon 1

- Series description: AC CT
- Slice: 3.00
- Kernel: "B18f very smooth"
- Window: "Abdomen"
- Extended FoV: On
- FoV: 700
- Center X: 0
- Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 111

3 Pause

4 PET WB

4.1 Routine

- Isotope: F-18
- Pharm.: FDG
- Inj. Dose: 10 Millicuries
- Inj. date (date/month - year): 30/10 - 2006
- Inj. time: 09:08:00
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 10 PtMinute

4.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram

- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

4.3 Recon

4.3.1 Recon 1

- Series description: PET Cardiac
- Recon range (bed): ups..
- Output image type: Corrected
- Recon method: Backprojection
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 2
- Filter: Gaussian
- FWHM (mm): 10
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off