# TUMOR\_HUND\_FDG.MlAdult

23. oktober 2012

# Indhold

1	Topogram 1			
	1.1	Routine	1	
	1.2	Scan	1	
<b>2</b>	CT	WB	1	
	2.1	Routine	1	
	2.2	Scan	2	
	2.3	Recons	2	
		2.3.1 Recon 1	2	
		2.3.2 Recon 2	3	
3	Pau	se	4	
4	$\mathbf{PE}^{r}$	T TUMOR	4	
_	4.1		4	
	4.2		4	
	4.3		5	
			5	
		4.3.2 Recon 2	5	
5	Pau	${f se}$	6	
6	$\mathbf{PE}^{r}$	$\Gamma$ WB	6	
Ü	6.1		6	
	6.2		7	
	6.3		7	
	0.0		7	
			8	
7	Pau	${f se}$	8	
8	Dvr	$\mathbf{Multi}$	8	
-	8.1		8	
	8.2		9	
	8.3	Recons	9	
		8 3 1 Recon 1	O	

# 1 Topogram

## 1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 0 mm
- Tube position: Top

### 1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 270 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: 80s shar
- Window: Topogram Body

# 2 CT WB

### 2.1 Routine

- Eff. mAs: 170
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDlvol: 11.4448mGy

• Scan time: 12.680 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 170 grader

# 2.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 170

• kV: 120

 $\bullet$  Scan time: 12.680 s

• Rotation time: 0.500 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 1.2000000

• Direction: Craniocaudal

### 2.3 Recons

### 2.3.1 Recon 1

• Series description: AC CT

• Slice: 3.00

• Kernel: B19f PET very smooth

• Window: Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 274

#### 2.3.2 Recon 2

• Series description: CT WB 3.0 B30f

• Slice: 3.00

• Kernel: B30f medium smooth

• Window: Abdomen

• Extended FoV: Off

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 274

# 3 Pause

# 4 PET TUMOR

### 4.1 Routine

• Isotope: F-18

• Pharm.: FDG

• Inj. Dose: 200 MegaBequerels

 $\bullet$  Inj. date (date/month - year): 20/06 - 2011

• Inj. time: 17:13:00

• Scan mode: Sinogram

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 10 Minutes

### 4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

• Input trigger signal: None

• LLD (keV): 435

• ULD (keV): 650

### 4.3 Recons

#### 4.3.1 Recon 1

- Series description: PET TUMOR Image check
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

#### 4.3.2 Recon 2

- Series description: PET TUMOR True X 2.0 mm
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 3
- Subsets: 21

- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: Off
- Match CT slice location: On
- Save intermediate data: Off

## 5 Pause

## 6 PET WB

### 6.1 Routine

- Isotope: F-18
- Pharm.: FDG
- Inj. Dose: 200 MegaBequerels
- Inj. date (date/month year): 20/06 2011
- Inj. time: 17:13:00
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 2 Minutes

### 6.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

• Input trigger signal: None

• LLD (keV): 435

• ULD (keV): 650

### 6.3 Recons

#### 6.3.1 Recon 1

• Series description: PET WB Image check

• Recon range (bed): Automatic/All/dunno

• Output image type: Corrected

• Recon method: Iterative

• Iterations: 4

• Subsets: 8

• Image size: 128

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 4

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: On

• Match CT slice location: On

• Save intermediate data: Off

#### 6.3.2 Recon 2

- Series description: PET WB True X 2.0 mm
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 3
- Subsets: 21
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: Off
- Match CT slice location: On
- Save intermediate data: Off

## 7 Pause

# 8 DynMulti

### 8.1 Routine

- Eff. mAs: 170
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDose

 $\bullet$  CTDlvol: 1032.37mGy

• Scan time: 90.100 s

• Delay: 4.000 s

• Slice: 7.20 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 170 grader

## 8.2 Scan

• Quality ref. mAs: 100

• Eff. mAs: 170

• kV: 120

• Scan time: 90.100 s

• Rotation time: 1.000 s

• Delay: 4.000 s

• Slice: 7.20 mm

• Pitch: 1.0000000

• Direction: Craniocaudal

### 8.3 Recons

#### 8.3.1 Recon 1

• Series description: DynMulti 7.2 B30s

• Slice: 7.20

• Kernel: B30s medium smooth

• Window: Abdomen

• Extended FoV: Off

• FoV: 0

• Center X: -3

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 1.000

• No. of images: 360