

3_PET_AC_LM_FET_Brain_neuroperfusion.Adult_PET4

17. november 2012

Indhold

1	Topogram	1
1.1	Routine	1
1.2	Scan	1
2	Lavdosis CT	1
2.1	Routine	1
2.2	Scan	2
2.3	Recons	2
2.3.1	Recon 1	2
2.3.2	Recon 2	3
3	Pause	4
4	PET Brain LM	4
4.1	Routine	4
4.2	Scan	4
4.3	Recons	5
4.3.1	Recon 1	5
4.3.2	Recon 2	5
4.3.3	Recon 3	6
4.3.4	Recon 4	7
5	Pause	7
6	ReadOnly=1 ResID=892 DynMulti4D	7
6.1	Routine	7
6.2	Scan	8
6.3	Recons	8
6.3.1	Recon 1	8
6.3.2	Recon 2	9
6.3.3	Recon 3	10
6.3.4	Recon 4	10

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 256 mm
- Tube position: Lateral

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 256 mm
- Direction: Craniocaudal
- Tube position: Lateral
- API: API
- Kernel: 20
- Window: Topogram

2 Lavdosis CT

2.1 Routine

- Eff. mAs: 30
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDIvol: 5.124mGy

- Scan time: 26.200 s
- Delay: 2.000 s
- Slice: 3 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

2.2 Scan

- Quality ref. mAs: 350
- Eff. mAs: 30
- kV: 120
- Scan time: 26.200 s
- Rotation time: 1.000 s
- Delay: 2.000 s
- Slice: 3 mm
- Pitch: 1.50
- Direction: Caudocranial

2.3 Recons

2.3.1 Recon 1

- Series description: tmp Lavdosis CT RTD
- Slice: 5
- Kernel: H19s
- Window: Cerebrum
- Extended FoV: Off
- FoV: 300
- Center X: 0

- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 5.000
- No. of images: 45

2.3.2 Recon 2

- Series description: AC Lavdosis CT
- Slice: 3
- Kernel: H19s
- Window: Cerebrum
- Extended FoV: Off
- FoV: 300
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 3.000
- No. of images: 74

3 Pause

4 PET Brain LM

4.1 Routine

- Isotope: F-18
- Pharm.: FET
- Inj. Dose: 185000000.0 Bequerels
- Inj. date (date/month - year): 19/09 - 2012
- Inj. time: 12:52:18
- Scan mode: List mode
- Scan range: Match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 20.000000 Minutes

4.2 Scan

- Autoload: Off
- Rebinner LUT: N/A
- Scan output: List mode
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- ULD (keV): N/A

4.3 Recons

4.3.1 Recon 1

- Series description: FET stat 1 mm 4
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400
- Zoom: 2.500000
- Filter: Gaussian
- FWHM (mm): 1.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

4.3.2 Recon 2

- Series description: FET stat 5 mm 4
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12

- Image size: 400
- Zoom: 2.500000
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

4.3.3 Recon 3

- Series description: FET stat 4Fdyn 5 mm 4
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400
- Zoom: 2.500000
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

4.3.4 Recon 4

- Series description: FET STAT 25-35 5 mm
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400
- Zoom: 2.500000
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

5 Pause

6 ReadOnly=1 ResID=892 DynMulti4D

6.1 Routine

- Eff. mAs: 181
- kV: 80
- CARE Dose4D: Off
- CareDoseType: CareDose

- CTDIvol: 328.42mGy
- Scan time: 1.500 s
- Delay: 4.000 s
- Slice: 5 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

6.2 Scan

- Quality ref. mAs: 250
- Eff. mAs: 181
- kV: 80
- Scan time: 1.500 s
- Rotation time: 0.330 s
- Delay: 4.000 s
- Slice: 5 mm
- Pitch: 1.10
- Direction: Caudocranial

6.3 Recons

6.3.1 Recon 1

- Series description: tmp DynMulti4D RTD
- Slice: 4
- Kernel: H20f
- Window: Cerebrum
- Extended FoV: Off
- FoV: 200

- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 31.300
- No. of images: 90

6.3.2 Recon 2

- Series description: VPCT DynMulti4D 5.0 H20f
- Slice: 5
- Kernel: H20f
- Window: Cerebrum
- Extended FoV: Off
- FoV: 200
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 5.000
- No. of images: 390

6.3.3 Recon 3

- Series description: VPCT DynMulti4D 5.0 H20f
- Slice: 5
- Kernel: H20f
- Window: Cerebrum
- Extended FoV: Off
- FoV: 200
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 5.000
- No. of images: 390

6.3.4 Recon 4

- Series description: DynMulti4D 5.0 H20f
- Slice: 5
- Kernel: H20f
- Window: Cerebrum
- Extended FoV: Off
- FoV: 200
- Center X: 0
- Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 5.000
- No. of images: 390