

PETCT_GantryOffset.Adult_PET4

30. maj 2013

Indhold

1	Topogram	1
1.1	Routine	1
1.2	Scan	1
2	CT	1
2.1	Routine	1
2.2	Scan	1
2.3	Recons	2
2.3.1	Recon 1	2
2.3.2	Recon 2	2
3	Pause	2
4	PET Offset	2
4.1	Routine	2
4.2	Scan	3
4.3	Recons	3
4.3.1	Recon 1	3

1 Topogram

1.1 Routine

- mA: 36
- kV: 120
- Topogram length: 512 mm
- Tube position: Top

1.2 Scan

- mA: 36
- kV: 120
- Delay: 4s
- Topogram length: 512 mm
- Direction: Craniocaudal
- Tube position: Top
- API: API
- Kernel: 80
- Window: Topogram

2 CT

2.1 Routine

- Eff. mAs: 50
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDIvol: 3.65446mGy
- Scan time: 8.039 s
- Delay: 2.000 s
- Slice: 2 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

2.2 Scan

- Quality ref. mAs: 120
- Eff. mAs: 50
- kV: 120
- Scan time: 8.039 s
- Rotation time: 0.500 s
- Delay: 2.000 s
- Slice: 2 mm
- Pitch: 0.80

- Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

- Series description: CT RTD
- Slice: 4
- Kernel: B19f
- Window: Abdomen
- Extended FoV: Off
- FoV: 500
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 4.000
- No. of images: 56

2.3.2 Recon 2

- Series description: CT 2.0 B19f
- Slice: 2
- Kernel: B19f
- Window: Abdomen
- Extended FoV: Off
- FoV: 500
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 111

3 Pause

4 PET Offset

4.1 Routine

- Isotope: Ge-68

- Pharm.: Crossed Rods
- Inj. Dose: 59200000 Bequerels
- Scan mode: Sinogram
- Scan range: Do not match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 10000000.000000 Counts

4.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): n/a
- ULD (keV): n/a

4.3 Recons

4.3.1 Recon 1

- Series description: PET Gantry Offset
- Recon range (bed): 1 to 1
- Output image type: Uncorrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 8
- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: Off
- Save intermediate data: Off