$PETCT_QualityCheck.MlAdult_PET5$

17. november 2012

Indhold

1	Topogram				
	1.1	Routine	1		
	1.2	Scan	1		
2	\mathbf{CT}		1		
	2.1	Routine	1		
	2.2	Scan	2		
	2.3	Recons	2		
		2.3.1 Recon 1	2		
3	Pau	${f se}$	3		
4	PET	Γ \mathbf{QC}	3		
	4.1		3		
	4.2		4		
	4.3	Recons	4		
		4.3.1 Recon 1	4		

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 512 mm
- Tube position: Top

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 512 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: 80s shar
- Window: Topogram Body

2 CT

2.1 Routine

- Eff. mAs: 50
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDlvol: 3.37586mGy

• Scan time: 5.630 s

• Delay: 4.000 s

• Slice: 5.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

2.2 Scan

• Quality ref. mAs: 50

• Eff. mAs: 50

• kV: 120

• Scan time: 5.630 s

• Rotation time: 0.500 s

• Delay: 4.000 s

• Slice: 5.00 mm

• Pitch: 0.8000000

• Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

• Series description: AC CT CT 5.0

• Slice: 5.00

• Kernel: B19f PET very smooth

• Window: Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 74

3 Pause

4 PET QC

4.1 Routine

• Isotope: Ge-68

• Pharm.: Solid Solution

• Inj. Dose: 93.15 MegaBequerels

• Inj. date (date/month - year): 20/01 - 2011

• Inj. time: 12:00:00

• Scan mode: Sinogram

• Scan range: Do not match CT FOV

• No. of beds: 1

• Scan duration/bed: 2e+008 Counts

4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

• Input trigger signal: None

• LLD (keV): 435

• ULD (keV): 650

4.3 Recons

4.3.1 Recon 1

• Series description: Norm Review

• Recon range (bed): Automatic/All/dunno

• Output image type: Corrected

• Recon method: Iterative

• Iterations: 2

• Subsets: 8

• Image size: 128

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 2.5

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: On

• Match CT slice location: Off

• Save intermediate data: Off