# $PET\_Cardiac\_FDG\_Projekt.MlAdult\_PET5$

27. februar 2013

## Indhold

1		ogram
	1.1	Routine
	1.2	Scan
2	$\mathbf{CT}$	
	2.1	Routine
	2.2	Scan
	2.3	Recons
		2.3.1 Recon 1
3	Pau	${f se}$
4	PET	$\Gamma  ext{ WB}$
	4.1	Routine
	4.2	Scan
	4.3	Recons
		4.3.1 Recon 1

## 1 Topogram

#### 1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 512 mm

• Tube position: Top

#### 1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 512 mm

• Direction: Craniocaudal

• Tube position: Top

• API: None

• Kernel: 80s shar

• Window: Topogram Body

## 2 CT WB

#### 2.1 Routine

• Eff. mAs: 40

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

• CTDlvol: 2.69288mGy

• Scan time: 12.680 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

#### 2.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 40

• kV: 120

 $\bullet$  Scan time: 12.680 s

• Rotation time: 0.500 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 1.2000000

• Direction: Craniocaudal

#### 2.3 Recons

#### 2.3.1 Recon 1

• Series description: AC CT

• Slice: 3.00

• Kernel: B19f PET very smooth

• Window: Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 274

#### 3 Pause

#### 4 PET WB

#### 4.1 Routine

• Isotope: F-18

• Pharm.: FDG

• Inj. Dose: 400 MegaBequerels

• Scan mode: List mode

• Scan range: Match CT FOV

• No. of beds: 0

• Scan duration/bed: 150 Seconds

#### 4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: List mode

• Sinogram mode: Trues

- Input trigger signal: ECG
- LLD (keV): 435
- ULD (keV): 650

#### 4.3 Recons

#### 4.3.1 Recon 1

- Series description: PET WB Uncorrected
- Recon range (bed): -1 to -1
- Output image type: NoRecon
- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: Off
- Match CT slice location: On
- Save intermediate data: Off