$4_4_PET_CT_WB_FDG_Terapi_Barn.Child$

8. oktober 2012

Indhold

1	Top	r -0														1									
	1.1	Routin	ne																						1
	1.2	Scan															•			•	•	•	•	•	1
2	CT	WB																							1
	2.1	Routin	ne																						1
	2.2	Scan																							2
	2.3	Recon																							2
		2.3.1	Recon 1																						2
		2.3.2	Recon 2																						3
		2.3.3	Recon 3																						4
		2.3.4	Recon 4																						4
		2.3.5	Recon 5																						5
3	Pau	se																							6
4	PET WB														6										
	4.1	Routin	ne																						6
	4.2	Scan																							6
	4.3	Recon																							7
		4.3.1	Recon 1																						7
		4.3.2	Recon 2																						7

1 Topogram

1.1 Routine

- mA: 30
- kV: 120
- Topogram length: 768 mm
- Tube position: Bottom

1.2 Scan

- mA: 30
- kV: 120
- Delay: 4s
- Topogram length: 768 mm
- Direction: Craniocaudal
- Tube position: Bottom
- API:
- Kernel: T80s
- Window: Topogra

2 CT WB

2.1 Routine

- Eff. mAs: 65
- kV: 120
- CARE Dose4D: On
- CTDlvol: 4.732mGy
- Scan time: 11.070 s

• Delay: 6.000 s

• Slice: 2 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

2.2 Scan

• Quality ref. mAs: 65

• Eff. mAs: 65

• kV: 120

• Scan time: 11.070 s

 \bullet Rotation time: 0.500 s

• Delay: 6.000 s

• Slice: 2 mm

• Pitch: 1.20

• Direction: Craniocaudal

2.3 Recon

2.3.1 Recon 1

• Series description: tmp CT WB RTD

• Slice: 4

• Kernel: "B19f"

• Window: "Abdomen"

• Extended FoV: false

• FoV: 500

• Center X: 0

• Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 4.000
- No. of images: 119

2.3.2 Recon 2

- Series description: AC CT WB
- Slice: 2
- Kernel: "B19f"
- Window: "Abdomen"
- Extended FoV: true
- FoV: 780
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 237

2.3.3 Recon 3

• Series description: CT WB 2.0 B30f

• Slice: 2

• Kernel: "B30f"

• Window: "Abdomen"

• Extended FoV: false

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 2.000

• No. of images: 237

2.3.4 Recon 4

• Series description: CT WB 2.0 SPO cor

• Slice: 2

• Kernel: "B30f"

• Window: "Abdomen"

• Extended FoV: false

• FoV: 500

• Center X: 0

• Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: 3D
- Recon Axis: Coronal
- Image order: AP
- Recon increment: 2.000
- No. of images: 251

2.3.5 Recon 5

- Series description: CT WB 2.0 SPO sag
- Slice: 2
- Kernel: "B30f"
- Window: "Abdomen"
- Extended FoV: false
- FoV: 500
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: 3D
- Recon Axis: Sagittal
- ullet Image order: LeftToRight
- Recon increment: 2.000
- No. of images: 251

3 Pause

4 PET WB

4.1 Routine

- Isotope: F-18
- Pharm.: FDG
- Inj. Dose: 1
- Inj. date (date/month year): 01/04 2011
- Inj. time: 14:35:47
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: N/A
- Scan duration/bed: 2.000000 Minutes

4.2 Scan

- Autoload: true
- Rebinner LUT: N/A
- Scan output: Sinogram
- Sinogram mode: PromptsRandoms
- Input trigger signal: None
- LLD (keV): N/A
- ULD (keV): N/A

4.3 Recon

4.3.1 Recon 1

- Series description: PET WB image check
- Recon range (bed): ups..
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 2
- Subsets: 12
- Image size: 200
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 3.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false

4.3.2 Recon 2

- Series description: PET WB Corr-TrueX
- Recon range (bed): ups..
- Output image type: Corrected
- Recon method: TrueXTOF
- Iterations: 2
- Subsets: 21

• Image size: 400

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 2.000000

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: true

• Match CT slice location: true

• Save intermediate data: false