

3_PET_AC_LM_FET_Brain_HGREB.Adult_PET4

3. december 2012

Indhold

1	Topogram	1
1.1	Routine	1
1.2	Scan	1
2	Lavdosis CT	1
2.1	Routine	1
2.2	Scan	2
2.3	Recons	2
2.3.1	Recon 1	2
2.3.2	Recon 2	3
3	Pause	4
4	PET Brain LM	4
4.1	Routine	4
4.2	Scan	4
4.3	Recons	5
4.3.1	Recon 1	5
4.3.2	Recon 2	5

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 256 mm
- Tube position: Lateral

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 256 mm
- Direction: Craniocaudal
- Tube position: Lateral
- API: API
- Kernel: 20
- Window: Topogram

2 Lavdosis CT

2.1 Routine

- Eff. mAs: 30
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDIvol: 5.124mGy

- Scan time: 26.200 s
- Delay: 2.000 s
- Slice: 3 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

2.2 Scan

- Quality ref. mAs: 350
- Eff. mAs: 30
- kV: 120
- Scan time: 26.200 s
- Rotation time: 1.000 s
- Delay: 2.000 s
- Slice: 3 mm
- Pitch: 1.50
- Direction: Caudocranial

2.3 Recons

2.3.1 Recon 1

- Series description: tmp Lavdosis CT RTD
- Slice: 5
- Kernel: H19s
- Window: Cerebrum
- Extended FoV: Off
- FoV: 300
- Center X: 0

- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 5.000
- No. of images: 45

2.3.2 Recon 2

- Series description: AC Lavdosis CT
- Slice: 3
- Kernel: H19s
- Window: Cerebrum
- Extended FoV: Off
- FoV: 300
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 3.000
- No. of images: 74

3 Pause

4 PET Brain LM

4.1 Routine

- Isotope: F-18
- Pharm.: FET
- Inj. Dose: 237000000.000000 Bequerels
- Inj. date (date/month - year): 02/04 - 2012
- Inj. time: 11:42:00
- Scan mode: List mode
- Scan range: Match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 20.000000 Minutes

4.2 Scan

- Autoload: Off
- Rebinner LUT: N/A
- Scan output: List mode
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- ULD (keV): N/A

4.3 Recons

4.3.1 Recon 1

- Series description: FET stat 1 mm
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 1.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

4.3.2 Recon 2

- Series description: FET stat 5 mm4
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12

- Image size: 400
- Zoom: 2.500000
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off