# $test\_WB\_FDG\_PET\_CT\_IV\_Lunge\_6.MlAdult\_PET6$

7. marts 2013

## Indhold

1	Topogram 1			
	1.1	Routine	1	
	1.2	Scan	1	
<b>2</b>	$\mathbf{CT}$	BANAN	1	
	2.1	Routine	1	
	2.2	Scan	1	
	2.3	Recons	2	
		2.3.1 Recon 1	2	
		2.3.2 Recon 2	2	
3	Pau	ise	2	
4	CT	WB	2	
	4.1	Routine	2	
	4.2	Scan	3	
	4.3	Recons	3	
		4.3.1 Recon 1	3	
		4.3.2 Recon 2	3	
		4.3.3 Recon 3	4	
		4.3.4 Recon 4	4	
5	Pau	ise	5	
6	PE	$\Gamma$ WB	5	
	6.1	Routine	5	
	6.2	Scan	5	
	6.3	Recons	5	
		6.3.1 Recon 1	5	
		6.3.2 Recon 2	6	

## 1 Topogram

#### 1.1 Routine

mA: 35kV: 120

 $\bullet$  Topogram length: 0 mm

• Tube position: Top

#### 1.2 Scan

mA: 35kV: 120Delay: 5s

Topogram length: 0 mmDirection: Craniocaudal

• Tube position: Top

API: API 31Kernel: 80s shar

• Window: Topogram Body

## 2 CT BANAN

#### 2.1 Routine

• Eff. mAs: 100

• kV: 120

• CARE Dose4D: On

• CareDoseType: CareDoseAEC

CTDlvol: 7.6543mGyScan time: 7.840 s

Delay: 8.000 sSlice: 5.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

#### 2.2 Scan

• Quality ref. mAs: 100

• Eff. mAs: 100

• kV: 120

Scan time: 7.840 sRotation time: 0.500 s

Delay: 8.000 sSlice: 5.00 mmPitch: 1.4000000

• Direction: Caudocranial

#### 2.3 Recons

#### 2.3.1 Recon 1

- Series description: CT BANAN B18f very smooth
- Slice: 5.00
- Kernel: B18f very smooth Window: Mediastinum
- Extended FoV: Off
- FoV: 315Center X: 0Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: AxialRecon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 127

#### 2.3.2 Recon 2

- Series description: CT BANAN 2.0 B70f
- Slice: 2.00
- Kernel: B70f very sharp
- $\bullet~$  Window: Lung
- Extended FoV: Off
- FoV: 315
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 1.500
- No. of images: 253

#### 3 Pause

### 4 CT WB

#### 4.1 Routine

• Eff. mAs: 225

• kV: 120

• CARE Dose4D: On

• CareDoseType: CareDoseAEC

CTDlvol: 15.1475mGyScan time: 12.680 s

Delay: 60.000 sSlice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

#### 4.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 225

• kV: 120

Scan time: 12.680 sRotation time: 0.500 s

Delay: 60.000 sSlice: 3.00 mmPitch: 1.2000000

• Direction: Craniocaudal

#### 4.3 Recons

#### 4.3.1 Recon 1

• Series description: AC CT

• Slice: 3.00

• Kernel: B18f very smooth

Window: AbdomenExtended FoV: On

FoV: 700Center X: 0Center Y: 0Mirroring: None

• Extended CT scale: Standard

Recon job: AxialRecon Axis: Axial

Image order: CraniocaudalRecon increment: 3.000

• No. of images: 274

#### 4.3.2 Recon 2

• Series description: CT WB 3.0 B30f

• Slice: 3.00

• Kernel: B30f medium smooth

Window: AbdomenExtended FoV: Off

FoV: 500Center X: 0Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

Recon job: AxialRecon Axis: Axial

Image order: CraniocaudalRecon increment: 3.000

• No. of images: 274

#### 4.3.3 Recon 3

• Series description: CT WB 3.0 COR

• Slice: 3.00

• Kernel: B30f medium smooth

Window: AbdomenExtended FoV: Off

FoV: 821Center X: 0Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: 3D

• Recon Axis: Coronal

• Image order: AP

• Recon increment: 3.000

• No. of images: 167

#### 4.3.4 Recon 4

• Series description: CT WB 3.0 SAG

• Slice: 3.00

• Kernel: B30f medium smooth

Window: AbdomenExtended FoV: Off

FoV: 821Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: 3D

- Recon Axis: Sagittal
- ullet Image order: LeftToRight
- Recon increment: 3.000
- No. of images: 274

## 5 Pause

## 6 PET WB

#### 6.1 Routine

- Isotope: F-18
- Pharm.: FDG
- Inj. Dose: 1 MegaBequerels
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 150 Seconds

#### 6.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

#### 6.3 Recons

#### 6.3.1 Recon 1

- Series description: PET WB Image check
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off (1)

- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

#### 6.3.2 Recon 2

- $\bullet$  Series description: PET WB True X 2 mm
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 3
- Subsets: 21
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off