

PETCT_NEMA_IQ.MlAdult_PET6

20. februar 2013

Indhold

1	Topogram	1
1.1	Routine	1
1.2	Scan	1
2	Topogram	1
2.1	Routine	1
2.2	Scan	2
3	CT Body	2
3.1	Routine	2
3.2	Scan	3
3.3	Recons	3
3.3.1	Recon 1	3
3.3.2	Recon 2	4
4	Pause	4
5	PET Control	4
5.1	Routine	4
5.2	Scan	5
5.3	Recons	5
5.3.1	Recon 1	5
6	Pause	6
7	PET LM1 13	6
7.1	Routine	6
7.2	Scan	6
7.3	Recons	7
7.3.1	Recon 1	7
8	Pause	8
9	PET LM2 14	8
9.1	Routine	8
9.2	Scan	8
9.3	Recons	8
9.3.1	Recon 1	8
10	Pause	9

11 PET LM3	15	9
11.1 Routine		9
11.2 Scan		10
11.3 Recons		10
11.3.1 Recon 1		10

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 512 mm
- Tube position: Top

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 512 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: 80s shar
- Window: Topogram Body

2 Topogram

2.1 Routine

- mA: 35
- kV: 120
- Topogram length: 512 mm
- Tube position: Lateral

2.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 512 mm
- Direction: Caudocranial
- Tube position: Lateral
- API: None
- Kernel: 80s shar
- Window: Topogram Body

3 CT Body

3.1 Routine

- Eff. mAs: 170
- kV: 120
- CARE Dose4D: On
- CareDoseType: CareDoseAEC
- CTDIvol: 11.4779mGy
- Scan time: 5.630 s
- Delay: 4.000 s
- Slice: 2.00 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

3.2 Scan

- Quality ref. mAs: 170
- Eff. mAs: 170
- kV: 120
- Scan time: 5.630 s
- Rotation time: 0.500 s
- Delay: 4.000 s
- Slice: 2.00 mm
- Pitch: 0.8000000
- Direction: Craniocaudal

3.3 Recons

3.3.1 Recon 1

- Series description: AC CT
- Slice: 2.00
- Kernel: B19f PET very smooth
- Window: Abdomen
- Extended FoV: On
- FoV: 700
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal

- Recon increment: 2.000
- No. of images: 111

3.3.2 Recon 2

- Series description: CT Body 2.0 B30f
- Slice: 2.00
- Kernel: B30f medium smooth
- Window: Abdomen
- Extended FoV: Off
- FoV: 500
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 111

4 Pause

5 PET Control

5.1 Routine

- Isotope: F-18
- Pharm.: FDG

- Inj. Dose: 10 Millicuries
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 60 Seconds

5.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

5.3 Recons

5.3.1 Recon 1

- Series description: PET WB
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian

- FWHM (mm): 5
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

6 Pause

7 PET LM1 13

7.1 Routine

- Isotope: F-18
- Pharm.: FDG
- Inj. Dose: 10 Millicuries
- Scan mode: List mode
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 13 Minutes

7.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: List mode
- Sinogram mode: Trues
- Input trigger signal: None

- LLD (keV): 435
- ULD (keV): 650

7.3 Recons

7.3.1 Recon 1

- Series description: PET WB
- Recon range (bed): -1 to -1
- Output image type: NoRecon
- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

8 Pause

9 PET LM2 14

9.1 Routine

- Isotope: F-18
- Pharm.: FDG
- Inj. Dose: 10 Millicuries
- Scan mode: List mode
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 14 Minutes

9.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: List mode
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

9.3 Recons

9.3.1 Recon 1

- Series description: PET WB
- Recon range (bed): -1 to -1
- Output image type: NoRecon

- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

10 Pause

11 PET LM3 15

11.1 Routine

- Isotope: F-18
- Pharm.: FDG
- Inj. Dose: 10 Millicuries
- Scan mode: List mode
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 15 Minutes

11.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: List mode
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

11.3 Recons

11.3.1 Recon 1

- Series description: PET WB
- Recon range (bed): -1 to -1
- Output image type: NoRecon
- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off