# $PETCT\_WB\_FDG\_IV\_45\_55\_kg\_Child.MlChild\_PET5$

3. december 2012

## Indhold

1	Topogram 1			
	1.1	Routine	1	
	1.2	Scan	1	
<b>2</b>	$\mathbf{CT}$		1	
	2.1	Routine	1	
	2.2	Scan	2	
	2.3	Recons	2	
		2.3.1 Recon 1	2	
		2.3.2 Recon 2	3	
3	Pau	ase	4	
4	PET	$\Gamma$ WB	4	
	4.1	Routine	4	
	4.2	Scan	4	
	4.3	Recons	5	
		4.3.1 Recon 1	5	
		4.3.2 Recon 2	-	

## 1 Topogram

## 1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 0 mm
- Tube position: Top

#### 1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 0 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: 20s standar
- Window: Topogram Body

## 2 CT

## 2.1 Routine

- Eff. mAs: 200
- kV: 120
- CARE Dose4D: On
- CareDoseType: CareDoseAEC
- CTDlvol: 13.4867mGy

• Scan time: 5.410 s

• Delay: 20.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

## 2.2 Scan

• Quality ref. mAs: 200

• Eff. mAs: 200

• kV: 120

 $\bullet$  Scan time: 5.410 s

• Rotation time: 0.500 s

• Delay: 20.000 s

• Slice: 3.00 mm

• Pitch: 1.4000000

• Direction: Craniocaudal

## 2.3 Recons

#### 2.3.1 Recon 1

• Series description: AC CT

• Slice: 3.00

• Kernel: B19f PET very smooth

• Window: Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 124

#### 2.3.2 Recon 2

• Series description: CT WB 3.0 B30f

• Slice: 3.00

• Kernel: B30f medium smooth

• Window: Abdomen

• Extended FoV: Off

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 124

## 3 Pause

## 4 PET WB

## 4.1 Routine

• Isotope: F-18

• Pharm.: FDG

• Inj. Dose: 1 MegaBequerels

 $\bullet\,$  Inj. date (date/month - year): 04/06 - 2012

• Inj. time: 17:16:00

• Scan mode: Sinogram

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 3 Minutes

#### 4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

• Input trigger signal: None

• LLD (keV): 425

• ULD (keV): 650

#### 4.3 Recons

#### 4.3.1 Recon 1

- Series description: PET WB Image check
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

#### 4.3.2 Recon 2

- Series description: PET WB Corrected True X 2 mm
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 2
- Subsets: 21

• Image size: 336

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 2

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: On

• Match CT slice location: On

• Save intermediate data: Off