

1\_PET\_CT\_WB\_FDG\_resp\_IV.Adult\_PET4

11. januar 2013

# Indhold

<b>1</b>	<b>Topogram</b>	<b>1</b>
1.1	Routine . . . . .	1
1.2	Scan . . . . .	1
<b>2</b>	<b>Lunge</b>	<b>1</b>
2.1	Routine . . . . .	1
2.2	Scan . . . . .	2
2.3	Recons . . . . .	2
2.3.1	Recon 1 . . . . .	2
2.3.2	Recon 2 . . . . .	3
2.3.3	Recon 3 . . . . .	4
<b>3</b>	<b>Pause</b>	<b>4</b>
<b>4</b>	<b>PreMonitoring</b>	<b>4</b>
4.1	Routine . . . . .	4
4.2	Scan . . . . .	5
4.3	Recons . . . . .	5
4.3.1	Recon 1 . . . . .	5
<b>5</b>	<b>Monitoring</b>	<b>6</b>
5.1	Routine . . . . .	6
5.2	Scan . . . . .	7
5.3	Recons . . . . .	7
5.3.1	Recon 1 . . . . .	7
<b>6</b>	<b>CT WB</b>	<b>8</b>
6.1	Routine . . . . .	8
6.2	Scan . . . . .	8
6.3	Recons . . . . .	9
6.3.1	Recon 1 . . . . .	9
6.3.2	Recon 2 . . . . .	9
6.3.3	Recon 3 . . . . .	10
6.3.4	Recon 4 . . . . .	11
6.3.5	Recon 5 . . . . .	11
<b>7</b>	<b>Pause</b>	<b>12</b>
<b>8</b>	<b>PET WB</b>	<b>12</b>
8.1	Routine . . . . .	12
8.2	Scan . . . . .	12
8.3	Recons . . . . .	13

8.3.1	Recon 1 . . . . .	13
8.3.2	Recon 2 . . . . .	14

# **1 Topogram**

## **1.1 Routine**

- mA: 35
- kV: 120
- Topogram length: 1536 mm
- Tube position: Top

## **1.2 Scan**

- mA: 35
- kV: 120
- Delay: 6s
- Topogram length: 1536 mm
- Direction: Craniocaudal
- Tube position: Top
- API: API 962494eb-db00-4e54-b01b-9e42ecd67603
- Kernel: 20
- Window: Topogram

# **2 Lunge**

## **2.1 Routine**

- Eff. mAs: 100
- kV: 120
- CARE Dose4D: On
- CareDoseType: CareDoseAEC
- CTDIvol: 7.6mGy

- Scan time: 7.210 s
- Delay: 9.000 s
- Slice: 5 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

## **2.2 Scan**

- Quality ref. mAs: 100
- Eff. mAs: 100
- kV: 120
- Scan time: 7.210 s
- Rotation time: 0.500 s
- Delay: 9.000 s
- Slice: 5 mm
- Pitch: 1.40
- Direction: Caudocranial

## **2.3 Recons**

### **2.3.1 Recon 1**

- Series description: tmp Lunge RTD
- Slice: 5
- Kernel: B18f
- Window: Mediastinum
- Extended FoV: Off
- FoV: 331
- Center X: 0

- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 5.000
- No. of images: 69

### **2.3.2 Recon 2**

- Series description: CT Lunge 5.0 B18f
- Slice: 5
- Kernel: B18f
- Window: Mediastinum
- Extended FoV: Off
- FoV: 331
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 115

### **2.3.3 Recon 3**

- Series description: CT Lunge 2.0 B70f
- Slice: 2
- Kernel: B70f
- Window: Lung
- Extended FoV: Off
- FoV: 331
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 1.500
- No. of images: 230

## **3 Pause**

## **4 PreMonitoring**

### **4.1 Routine**

- Eff. mAs: 20
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDIvol: 1.186mGy

- Scan time: 0.500 s
- Delay: 2.000 s
- Slice: 10 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

## **4.2 Scan**

- Quality ref. mAs: 225
- Eff. mAs: 20
- kV: 120
- Scan time: 0.500 s
- Rotation time: 0.500 s
- Delay: 2.000 s
- Slice: 10 mm
- Pitch: 0.00
- Direction: Craniocaudal

## **4.3 Recons**

### **4.3.1 Recon 1**

- Series description: PreMonitoring 10.0 eFoV
- Slice: 10
- Kernel: B30s
- Window: Abdomen
- Extended FoV: On
- FoV: 780
- Center X: 0



- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 10.000
- No. of images: 1

## 5 Monitoring

### 5.1 Routine

- Eff. mAs: 20
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDIvol: 35.58mGy
- Scan time: 0.500 s
- Delay: 10.000 s
- Slice: 10 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

## **5.2 Scan**

- Quality ref. mAs: 225
- Eff. mAs: 20
- kV: 120
- Scan time: 0.500 s
- Rotation time: 0.500 s
- Delay: 10.000 s
- Slice: 10 mm
- Pitch: 0.00
- Direction: Craniocaudal

## **5.3 Recons**

### **5.3.1 Recon 1**

- Series description: Monitoring 10.0 eFoV
- Slice: 10
- Kernel: B30s
- Window: Abdomen
- Extended FoV: On
- FoV: 780
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal

- Recon increment: 10.000
- No. of images: 30

## **6 CT WB**

### **6.1 Routine**

- Eff. mAs: 225
- kV: 120
- CARE Dose4D: On
- CareDoseType: CareDoseAEC
- CTDIvol: 16.408mGy
- Scan time: 34.870 s
- Delay: 20.000 s
- Slice: 2 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

### **6.2 Scan**

- Quality ref. mAs: 225
- Eff. mAs: 225
- kV: 120
- Scan time: 34.870 s
- Rotation time: 0.500 s
- Delay: 20.000 s
- Slice: 2 mm
- Pitch: 0.65
- Direction: Craniocaudal

## **6.3 Recons**

### **6.3.1 Recon 1**

- Series description: tmp CT WB RTD
- Slice: 4
- Kernel: B30f
- Window: Abdomen
- Extended FoV: Off
- FoV: 500
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 4.000
- No. of images: 213

### **6.3.2 Recon 2**

- Series description: CT WB 2.0 eFoV
- Slice: 2
- Kernel: B30f
- Window: Abdomen
- Extended FoV: On
- FoV: 780
- Center X: 0

- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 426

### **6.3.3 Recon 3**

- Series description: CT WB 2.0 B30f
- Slice: 2
- Kernel: B30f
- Window: Abdomen
- Extended FoV: Off
- FoV: 500
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 426

#### **6.3.4 Recon 4**

- Series description: CT WB 2.0 cor B30f
- Slice: 2
- Kernel: B30f
- Window: Abdomen
- Extended FoV: Off
- FoV: 977
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: 3D
- Recon Axis: Coronal
- Image order: AP
- Recon increment: 2.000
- No. of images: 151

#### **6.3.5 Recon 5**

- Series description: CT WB 2.0 sag B30f
- Slice: 2
- Kernel: B30f
- Window: Abdomen
- Extended FoV: Off
- FoV: 977
- Center X: 0
- Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: 3D
- Recon Axis: Sagittal
- Image order: LeftToRight
- Recon increment: 2.000
- No. of images: 209

## **7 Pause**

## **8 PET WB**

### **8.1 Routine**

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- Inj. Dose: 1800000000 Bequerels
- Inj. date (date/month - year): 08/06 - 2012
- Inj. time: 11:15:00
- Scan mode: Sinogram
- Scan range:
- No. of beds: Not given. Check recon range
- Scan duration/bed: 2.000000 Minutes

### **8.2 Scan**

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram

- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): n/a
- ULD (keV): n/a

## **8.3 Recons**

### **8.3.1 Recon 1**

- Series description: PET WB Image Check HD
- Recon range (bed): 1 to 6
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 2
- Subsets: 12
- Image size: 200
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off



### 8.3.2 Recon 2

- Series description: PET WB TrueX + TOF 2 mm
- Recon range (bed): 1 to 6
- Output image type: Corrected
- Recon method: TrueXTOF
- Iterations: 2
- Subsets: 21
- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off