# PETCT\_NEMA\_IQ.MlAdult\_PET6

30. maj 2013

## Indhold

1	_	ogram	1		
	1.1	Routine	1		
	1.2	Scan	1		
2	Topogram				
	2.1	Routine	1		
	2.2	Scan	1		
3	CT	Body	1		
	3.1	Routine	1		
	3.2	Scan	2		
	3.3	Recons	2		
		3.3.1 Recon 1	2		
		3.3.2 Recon 2	2		
4	Pau	${f se}$	3		
5	PET Control				
	5.1	Routine	3		
	5.2	Scan	3		
	5.3	Recons	3		
		5.3.1 Recon 1	3		
6	Pau	se	4		
7	PET LM1 13				
	7.1	Routine	4		
	7.2	Scan	4		
	7.3	Recons	4		
		7.3.1 Recon 1	4		
8	Pau	se	5		
9	PE	Γ LM2 14	5		
	9.1	Routine	5		
	9.2	Scan	5		
	9.3	Recons	5		
		9.3.1 Recon 1	5		
10	Pau	se	6		

11 PET LM3 15	6
11.1 Routine	. 6
11.2 Scan	. 6
11.3 Recons	. 6
11.3.1 Recon 1	. 6

## 1 Topogram

## 1.1 Routine

mA: 35kV: 120

• Topogram length: 512 mm

• Tube position: Top

## 1.2 Scan

mA: 35kV: 120Delay: 4s

Topogram length: 512 mmDirection: Craniocaudal

• Tube position: Top

• API: None

• Kernel: 80s shar

• Window: Topogram Body

## 2 Topogram

## 2.1 Routine

mA: 35kV: 120

• Topogram length: 512 mm

• Tube position: Lateral

## 2.2 Scan

mA: 35kV: 120Delay: 4s

Topogram length: 512 mm
Direction: Caudocranial
Tube position: Lateral

• Tube position: Lateral

• API: None

• Kernel: 80s shar

• Window: Topogram Body

## 3 CT Body

## 3.1 Routine

• Eff. mAs: 170

- kV: 120
- CARE Dose4D: On
- CareDoseType: CareDoseAEC
- CTDlvol: 11.4779mGy
- Scan time: 5.630 s
- Delay: 4.000 s
- Slice: 2.00 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

#### 3.2 Scan

- Quality ref. mAs: 170
- Eff. mAs: 170
- kV: 120
- Scan time: 5.630 s
- Rotation time: 0.500 s
- Delay: 4.000 s
- Slice: 2.00 mm
- Pitch: 0.8000000
- Direction: Craniocaudal

#### 3.3 Recons

#### 3.3.1 Recon 1

- Series description: AC CT
- Slice: 2.00
- Kernel: B19f PET very smooth
- Window: Abdomen
- Extended FoV: On
- FoV: 700
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 111

#### 3.3.2 Recon 2

- Series description: CT Body 2.0 B30f
- Slice: 2.00

• Kernel: B30f medium smooth

Window: AbdomenExtended FoV: Off

FoV: 500Center X: 0Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

Recon job: AxialRecon Axis: Axial

Image order: CraniocaudalRecon increment: 2.000No. of images: 111

## 4 Pause

## 5 PET Control

## 5.1 Routine

Isotope: F-18Pharm.: FDG

Inj. Dose: 10 MillicuriesScan mode: Sinogram

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 60 Seconds

## 5.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

 $\bullet\,$  Sinogram mode: Trues

• Input trigger signal: None

LLD (keV): 435ULD (keV): 650

## 5.3 Recons

#### 5.3.1 Recon 1

• Series description: PET WB

 $\bullet$  Recon range (bed): -1 to -1

• Output image type: Corrected

• Recon method: Iterative

- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

## 6 Pause

## 7 PET LM1 13

## 7.1 Routine

- Isotope: F-18Pharm.: FDG
- Inj. Dose: 10 Millicuries
- Scan mode: List mode
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 13 Minutes

#### 7.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: List mode
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

## 7.3 Recons

#### 7.3.1 Recon 1

- Series description: PET WB
- Recon range (bed): -1 to -1
- Output image type: NoRecon
- Recon method: Iterative
- Iterations: 4

- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

## 8 Pause

## 9 PET LM2 14

## 9.1 Routine

- Isotope: F-18
- Pharm.: FDG
- Inj. Dose: 10 Millicuries
- Scan mode: List mode
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 14 Minutes

## 9.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: List mode
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

#### 9.3 Recons

## 9.3.1 Recon 1

- Series description: PET WB
- Recon range (bed): -1 to -1
- Output image type: NoRecon
- Recon method: Iterative
- Iterations: 4
- Subsets: 8

- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

## 10 Pause

## 11 PET LM3 15

## 11.1 Routine

- Isotope: F-18
- Pharm.: FDG
- Inj. Dose: 10 Millicuries
- Scan mode: List mode
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 15 Minutes

#### 11.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: List mode
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

#### 11.3 Recons

#### 11.3.1 Recon 1

- Series description: PET WB
- Recon range (bed): -1 to -1
- Output image type: NoRecon
- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168

• Zoom: 1

Filter: GaussianFWHM (mm): 5Offset X: 0 mmOffset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: On

Match CT slice location: On Save intermediate data: Off