# $4_2$ PET\_AC\_WB\_FDG\_HER3.Adult

22. oktober 2012

# Indhold

1	Top	ogram	1		
	1.1	Routine	1		
	1.2	Scan	]		
<b>2</b>	LavdoseCTWB				
	2.1	Routine	1		
	2.2	Scan	2		
	2.3	Recons	2		
		2.3.1 Recon 1	2		
		2.3.2 Recon 2	•		
3	Pau	se	4		
4	PE	WB	4		
	4.1	Routine	4		
	4.2	Scan	4		
	4.3	Recons			
		4.3.1 Recon 1			
		4.3.2 Recon 2			
		4.3.3 Recon 3	6		

## 1 Topogram

## 1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 1024 mm
- Tube position: Top

#### 1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 1024 mm
- Direction: Craniocaudal
- Tube position: Top
- API: API
- Kernel: 80
- Window: Topogram

## 2 LavdoseCTWB

## 2.1 Routine

- Eff. mAs: 50
- kV: 120
- CARE Dose4D: On
- CareDoseType: CareDoseAEC
- CTDlvol: 3.64mGy

• Scan time: 32.530 s

• Delay: 2.000 s

• Slice: 2 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

## 2.2 Scan

• Quality ref. mAs: 50

• Eff. mAs: 50

• kV: 120

 $\bullet$  Scan time: 32.530 s

• Rotation time: 0.500 s

• Delay: 2.000 s

• Slice: 2 mm

• Pitch: 0.80

• Direction: Craniocaudal

### 2.3 Recons

#### 2.3.1 Recon 1

• Series description: tmp LavdoseCTWB RTD

• Slice: 4

• Kernel: B19f

• Window: Abdomen

• Extended FoV: false

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 4.000

• No. of images: 245

#### 2.3.2 Recon 2

• Series description: CT transmission scan

• Slice: 2

• Kernel: B19f

• Window: Abdomen

• Extended FoV: true

• FoV: 780

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 2.000

• No. of images: 489

## 3 Pause

## 4 PET WB

### 4.1 Routine

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- $\bullet\,$  Inj. Dose: 1000000.000000 Bequerels
- Inj. date (date/month year): 24/05 2012
- Inj. time: 16:14:43
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 2.000000 Minutes

#### 4.2 Scan

- Autoload: true
- Rebinner LUT: N/A
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- ULD (keV): N/A

#### 4.3 Recons

#### 4.3.1 Recon 1

- Series description: PET WB Image check HD
- Recon range (bed): 1 to 7
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 8
- Image size: 200
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false

#### 4.3.2 Recon 2

- Series description: PET AC
- Recon range (bed): 1 to 7
- Output image type: Corrected
- Recon method: TrueXTOF
- Iterations: 2
- Subsets: 21

• Image size: 400

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 2.000000

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: true

• Match CT slice location: true

• Save intermediate data: false

#### 4.3.3 Recon 3

• Series description: PET NAC

• Recon range (bed): 1 to 7

• Output image type: Uncorrected

• Recon method: Iterative3D

• Iterations: 2

• Subsets: 24

• Image size: 400

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 2.000000

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: true

• Match CT slice location: true

• Save intermediate data: false