# PET\_AC\_WB\_FDG.MlChild\_PET5

11. januar 2013

# Indhold

1	Top	Topogram															1						
	1.1	Routine																					1
	1.2	Scan .								•			•					•					1
<b>2</b>	CT lavdosis																						
	2.1	Routine																					1
	2.2	Scan .																					2
	2.3	Recons																					2
		2.3.1	Recon 1																				2
3	Pau	ıse																					3
4	PET WB														3								
	4.1	Routine																					3
	4.2	Scan .																					4
	4.3																						4
			Recon 1																				
		4.3.2	Recon 2																				5

# 1 Topogram

# 1.1 Routine

- mA: 50
- kV: 80
- Topogram length: 0 mm
- Tube position: Top

#### 1.2 Scan

- mA: 50
- kV: 80
- Delay: 4s
- Topogram length: 0 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: 20s standar
- Window: Topogram Body

# 2 CT lavdosis

## 2.1 Routine

- Eff. mAs: 30
- kV: 80
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- $\bullet$  CTDlvol: 0.540268mGy

• Scan time: 11.000 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

# 2.2 Scan

• Quality ref. mAs: 25

• Eff. mAs: 30

• kV: 80

 $\bullet$  Scan time: 11.000 s

• Rotation time: 0.500 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 1.4000000

• Direction: Craniocaudal

# 2.3 Recons

## 2.3.1 Recon 1

• Series description: AC CT lavdosis

• Slice: 3.00

• Kernel: B19f PET very smooth

• Window: Baby Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 274

# 3 Pause

# 4 PET WB

### 4.1 Routine

• Isotope: F-18

• Pharm.: FDG

• Inj. Dose: 1 MegaBequerels

• Inj. date (date/month - year): 08/05 - 2012

• Inj. time: 17:05:00

• Scan mode: Sinogram

• Scan range: Match CT FOV

• No. of beds: 5

• Scan duration/bed: 2 Minutes

### 4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

• Input trigger signal: None

• LLD (keV): 425

• ULD (keV): 650

## 4.3 Recons

#### 4.3.1 Recon 1

• Series description: PET WB corrected

• Recon range (bed): Automatic/All/dunno

• Output image type: Corrected

• Recon method: Iterative

• Iterations: 4

• Subsets: 8

• Image size: 168

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 4

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: On

• Match CT slice location: On

• Save intermediate data: Off

#### 4.3.2 Recon 2

- $\bullet$  Series description: PET WB True X
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 2
- Subsets: 21
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: Off
- Match CT slice location: On
- Save intermediate data: Off