# 6\_Fantom.MlAdult

22. oktober 2012

# Indhold

1	Top	ogram	1	
	1.1	Routine	1	
	1.2	Scan	1	
<b>2</b>	$\mathbf{CT}$	Uden spejl	1	
	2.1	Routine	1	
	2.2	Scan	2	
	2.3	Recons	2	
		2.3.1 Recon 1	2	
		2.3.2 Recon 2	3	
3	Pau	ise	4	
4	CT	med spejl	4	
	4.1	Routine	4	
	4.2		4	
	4.3	Recons	5	
		4.3.1 Recon 1	5	
		4.3.2 Recon 2	5	
5	Pau	ase	6	
6	PET med spejl			
	6.1	Routine	6	
	6.2	Scan	7	
	6.3	Recons	7	
		6.3.1 Recon 1	7	
		6.3.2 Recon 2	8	

## 1 Topogram

## 1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 256 mm

• Tube position: Top

#### 1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 270 mm

• Direction: Craniocaudal

• Tube position: Top

• API: None

• Kernel: 80s shar

• Window: Topogram Body

# 2 CT Uden spejl

### 2.1 Routine

• Eff. mAs: 225

• kV: 120

• CARE Dose4D: On

• CareDoseType: CareDoseAEC

• CTDlvol: 15.1475mGy

• Scan time: 4.000 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 225 grader

## 2.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 225

• kV: 120

 $\bullet$  Scan time: 4.000 s

• Rotation time: 0.500 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 1.2000000

• Direction: Craniocaudal

## 2.3 Recons

#### 2.3.1 Recon 1

• Series description: tmp CT Uden spejl RTD

• Slice: 5.00

• Kernel: B19f PET very smooth

• Window: Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 17.300

• No. of images: 14

#### 2.3.2 Recon 2

• Series description: AC CT Uden Spejl

• Slice: 3.00

• Kernel: B19f PET very smooth

• Window: Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 74

## 3 Pause

## 4 CT med spejl

### 4.1 Routine

• Eff. mAs: 225

• kV: 120

• CARE Dose4D: On

• CareDoseType: CareDoseAEC

• CTDlvol: 15.1475mGy

• Scan time: 4.000 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 225 grader

### 4.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 225

• kV: 120

• Scan time: 4.000 s

 $\bullet$  Rotation time: 0.500 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 1.2000000

• Direction: Craniocaudal

#### 4.3 Recons

#### 4.3.1 Recon 1

- Series description: tmp CT med spejl RTD
- Slice: 5.00
- Kernel: B19f PET very smooth
- Window: Abdomen
- Extended FoV: On
- FoV: 700
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 17.300
- No. of images: 14

#### 4.3.2 Recon 2

- Series description: AC CT Med Spejl
- Slice: 3.00
- Kernel: B19f PET very smooth
- Window: Abdomen
- Extended FoV: On
- FoV: 700
- Center X: 0

- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 74

## 5 Pause

## 6 PET med spejl

#### 6.1 Routine

- Isotope: F-18
- Pharm.: FDG
- Inj. Dose: 25 MegaBequerels
- Inj. date (date/month year): 12/07 2012
- Inj. time: 14:47:00
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 5 Minutes

#### 6.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

• Input trigger signal: None

• LLD (keV): 435

• ULD (keV): 650

#### 6.3 Recons

#### 6.3.1 Recon 1

• Series description: AC PET m spejl CT u spejl

• Recon range (bed): Automatic/All/dunno

• Output image type: Corrected

• Recon method: TrueX

• Iterations: 6

• Subsets: 16

• Image size: 336

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 2

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: On

• Match CT slice location: On

• Save intermediate data: Off

#### 6.3.2 Recon 2

- $\bullet\,$  Series description: AC PET m spejl CT m spejl
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 6
- Subsets: 16
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off