

PETCT_WB_FDG_IV_25_35_kg_Child.MlChild_PET5

7. marts 2013

Indhold

1	Topogram	1
1.1	Routine	1
1.2	Scan	1
2	CT	1
2.1	Routine	1
2.2	Scan	1
2.3	Recons	2
2.3.1	Recon 1	2
2.3.2	Recon 2	2
3	Pause	2
4	PET WB	2
4.1	Routine	2
4.2	Scan	3
4.3	Recons	3
4.3.1	Recon 1	3
4.3.2	Recon 2	3

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 0 mm
- Tube position: Top

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 0 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: 20s standar
- Window: Topogram Body

2 CT

2.1 Routine

- Eff. mAs: 60
- kV: 120
- CARE Dose4D: On
- CareDoseType: CareDoseAEC
- CTDIvol: 4.04601mGy
- Scan time: 9.140 s
- Delay: 20.000 s
- Slice: 3.00 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

2.2 Scan

- Quality ref. mAs: 60
- Eff. mAs: 60
- kV: 120
- Scan time: 9.140 s
- Rotation time: 0.500 s
- Delay: 20.000 s
- Slice: 3.00 mm
- Pitch: 1.4000000

- Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

- Series description: AC CT
- Slice: 3.00
- Kernel: B19f PET very smooth
- Window: Abdomen
- Extended FoV: On
- FoV: 700
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 224

2.3.2 Recon 2

- Series description: CT WB 3.0 B30f
- Slice: 3.00
- Kernel: B30f medium smooth
- Window: Abdomen
- Extended FoV: Off
- FoV: 400
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 224

3 Pause

4 PET WB

4.1 Routine

- Isotope: F-18

- Pharm.: FDG
- Inj. Dose: 1 MegaBequerels
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 3 Minutes

4.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 425
- ULD (keV): 650

4.3 Recons

4.3.1 Recon 1

- Series description: PET WB Image check
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

4.3.2 Recon 2

- Series description: PET WB Corrected True X 2 mm
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 2
- Subsets: 21

- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off