$3_PET_AC_LM_FET_Brain_Dyn.Adult_PET4$

7. marts 2013

Indhold

| 1 | Top | ogram | 1 |
|----------|-----|----------------|---|
| | 1.1 | Routine | 1 |
| | 1.2 | Scan | 1 |
| 2 | Lav | dosis CT | 1 |
| | 2.1 | Routine | 1 |
| | 2.2 | Scan | 1 |
| | 2.3 | Recons | 2 |
| | | 2.3.1 Recon 1 | 2 |
| | | 2.3.2 Recon 2 | 2 |
| 3 | Pau | ase | 2 |
| 4 | PE | Γ Brain LM | 2 |
| | 4.1 | Routine | 2 |
| | 4.2 | Scan | 3 |
| | 4.3 | Recons | 3 |
| | | 4.3.1 Recon 1 | 3 |
| | | 4.3.2 Recon 2 | 3 |
| | | 4.3.3 Recon 3 | 4 |
| | | 4.3.3 RECOIL 3 | 7 |

1 Topogram

1.1 Routine

mA: 35kV: 120

Topogram length: 256 mmTube position: Lateral

1.2 Scan

mA: 35kV: 120Delay: 4s

Topogram length: 256 mmDirection: CraniocaudalTube position: Lateral

API: APIKernel: 20

• Window: Topogram

2 Lavdosis CT

2.1 Routine

• Eff. mAs: 350

• kV: 120

• CARE Dose4D: On

• CareDoseType: CareDoseAEC

CTDlvol: 59.7805mGyScan time: 26.000 s

Delay: 2.000 sSlice: 3 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

2.2 Scan

• Quality ref. mAs: 350

• Eff. mAs: 350

• kV: 120

Scan time: 26.000 sRotation time: 1.000 s

Delay: 2.000 sSlice: 3 mmPitch: 1.50

• Direction: Caudocranial

2.3 Recons

2.3.1 Recon 1

• Series description: Lavdosis CT RTD

Slice: 1.5Kernel: H19s

Window: CerebrumExtended FoV: Off

FoV: 300Center X: 0Center Y: 0Mirroring: None

• Extended CT scale: Standard

Recon job: AxialRecon Axis: Axial

Image order: CaudocranialRecon increment: 1.500

• No. of images: 148

2.3.2 Recon 2

• Series description: AC Lavdosis CT

Slice: 3Kernel: H19s

Window: CerebrumExtended FoV: Off

FoV: 300Center X: 0Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

Recon job: AxialRecon Axis: Axial

Image order: CaudocranialRecon increment: 3.000

• No. of images: 74

3 Pause

4 PET Brain LM

4.1 Routine

• Isotope: F-18

- Pharm.: FET
- $\bullet\,$ Inj. Dose: 185000000 Bequerels
- Scan mode: List mode
- Scan range: Do not match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 40.000000 Minutes

4.2 Scan

- Autoload: Off
- Rebinner LUT: Off
- Scan output: List mode
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): n/a
- ULD (keV): n/a

4.3 Recons

4.3.1 Recon 1

- \bullet Series description: FET dyn 22F 5 mm 4
- Recon range (bed): 1 to 1
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400
- Zoom: 2.500000
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

4.3.2 Recon 2

- Series description: FET 20-40 min 5 mm 4
- Recon range (bed): 1 to 1
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12

- Image size: 400
- Zoom: 2.500000
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mmOffset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

4.3.3 Recon 3

- Series description: FET dyn 25-35min 5mm
- Recon range (bed): 1 to 1
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400
- Zoom: 2.500000
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

4.3.4 Recon 4

- Series description: FET dyn 10-30min 5mm
- Recon range (bed): 1 to 1
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400
- Zoom: 2.500000
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()

• Scatter correction: On

Match CT slice location: On Save intermediate data: Off