$PETCT_NEMA_2007_IQ.Adult_PET4$

17. november 2012

Indhold

1	Topogram				
	1.1	Routine	1		
	1.2	Scan	1		
2	\mathbf{AC}	CT NEMA	1		
	2.1	Routine	1		
	2.2	Scan	2		
	2.3	Recons	2		
		2.3.1 Recon 1	2		
		2.3.2 Recon 2	3		
3	Pau	se	4		
4	PE	NEMA IQ	4		
	4.1	Routine	4		
	4.2	Scan	4		
	4.3	Recons	5		
		4.3.1 Recon 1	5		

1 Topogram

1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 1024 mm

• Tube position: Top

1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 1024 mm

• Direction: Craniocaudal

• Tube position: Top

• API: API

• Kernel: 80

• Window: Topogram

2 AC CT NEMA

2.1 Routine

• Eff. mAs: 50

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

• CTDlvol: 3.64mGy

• Scan time: 8.050 s

• Delay: 2.000 s

• Slice: 5 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

2.2 Scan

• Quality ref. mAs: 120

• Eff. mAs: 50

• kV: 120

 \bullet Scan time: 8.050 s

• Rotation time: 0.500 s

• Delay: 2.000 s

• Slice: 5 mm

• Pitch: 0.80

• Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

• Series description: tmp AC CT NEMA RTD

• Slice: 4

• Kernel: B19f

• Window: Abdomen

• Extended FoV: Off

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 4.000

• No. of images: 56

2.3.2 Recon 2

• Series description: AC CT NEMA

• Slice: 5

• Kernel: B19f

• Window: Abdomen

• Extended FoV: On

• FoV: 780

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 75

3 Pause

4 PET NEMA IQ

4.1 Routine

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- Inj. Dose: 370000000.000000 Bequerels
- Inj. date (date/month year): 28/10 2009
- Inj. time: 14:49:29
- Scan mode: Sinogram
- Scan range: Do not match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 7.600000 Minutes

4.2 Scan

- Autoload: On
- Rebinner LUT: N/A
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- ULD (keV): N/A

4.3 Recons

4.3.1 Recon 1

- Series description: PET NEMA IQ
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 2
- Subsets: 24
- Image size: 200
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: Off
- Save intermediate data: Off