# $TUMOR\_HYPOXI\_HUND\_LM.MlAdult$

1. oktober 2012

# Indhold

1		ogram
	1.1	Routine
	1.2	Scan
2	$\mathbf{CT}$	
	2.1	Routine
	2.2	Scan
	2.3	Recon
		2.3.1 Recon 1
3	Pau	${f se}$
4	PE:	r WB
	4.1	Routine
	4.2	Scan
	4.3	Recon
		4.3.1 Recon 1

# 1 Topogram

#### 1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 256 mm

• Tube position: Lateral

#### 1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 256 mm

• Direction: Craniocaudal

• Tube position: Lateral

• API: None

• Kernel: T20s standard

• Window: Topogram Head

# 2 CT WB

## 2.1 Routine

• Eff. mAs: 40

• kV: 9

• CARE Dose4D: Off

• CTDlvol: 5.504mGy

 $\bullet$  Scan time: 1.0000000 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 40 grader

## 2.2 Scan

• Quality ref. mAs: 380

• Eff. mAs: 40

• kV: 120

• Scan time: 9.310 s

• Rotation time: 1.000 s

• Delay: 4.000 s

• Slice: 0.6 mm

• Pitch: 1.0000000

• Direction: Craniocaudal

## 2.3 Recon

#### 2.3.1 Recon 1

 $\bullet\,$  Series description: AC CT

• Slice: 3.00

• Kernel: "H19s PET very smooth"

• Window: "Cerebrum"

• Extended FoV: Off

• FoV: 300

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 74

# 3 Pause

# 4 PET WB

## 4.1 Routine

• Isotope: Cu-64

• Pharm.: ATSM

• Inj. Dose: 700 MegaBequerels

• Inj. date (date/month - year): 03/11 - 2009

• Inj. time: 15:18:00

• Scan mode: List mode

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 60 PtMinute

#### 4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: List mode

- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

## 4.3 Recon

#### 4.3.1 Recon 1

- Series description: PET Brain LM
- Recon range (bed): ups..
- Output image type: NoRecon
- Recon method: Iterative
- Iterations: 6
- Subsets: 16
- Image size: 256
- Zoom: 2
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off