

PETCT_Brain_Ped.MlChild

27. september 2012

Indhold

1	Topogram	1
1.1	Routine	1
1.2	Scan	1
2	CT WB	1
2.1	Routine	1
2.2	Scan	2
2.3	Recon	2
2.3.1	Recon 1	2
2.3.2	Recon 2	3
3	Pause	4
4	PET WB	4
4.1	Routine	4
4.2	Scan	4
4.3	Recon	5
4.3.1	Recon 1	5

1 Topogram

1.1 Routine

- mA: 50
- kV: 80
- Topogram length: 256 mm
- Tube position: Lateral

1.2 Scan

- mA: 50
- kV: 80
- Delay: 4s
- Topogram length: 256 mm
- Direction: Craniocaudal
- Tube position: Lateral
- API: None
- Kernel: T20s standard
- Window: Topogram Head

2 CT WB

2.1 Routine

- Eff. mAs: 170
- kV: 24
- CARE Dose4D: Off
- CTDIvol: 28.798mGy
- Scan time: 0.8000000 s

- Delay: 4.000 s
- Slice: 3.00 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 170 grader

2.2 Scan

- Quality ref. mAs: 380
- Eff. mAs: 170
- kV: 120
- Scan time: 24.750 s
- Rotation time: 1.000 s
- Delay: 4.000 s
- Slice: 0.6 mm
- Pitch: 0.8000000
- Direction: Craniocaudal

2.3 Recon

2.3.1 Recon 1

- Series description: AC CT
- Slice: 3.00
- Kernel: "H19s PET very smooth"
- Window: "Cerebrum"
- Extended FoV: Off
- FoV: 300
- Center X: 0
- Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 1.500
- No. of images: 148

2.3.2 Recon 2

- Series description: CT Brain 3.0 C30s
- Slice: 3.00
- Kernel: "C30s medium smooth"
- Window: "Baby Brain"
- Extended FoV: Off
- FoV: 250
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 1.500
- No. of images: 148

3 Pause

4 PET WB

4.1 Routine

- Isotope: F-18
- Pharm.: FDG
- Inj. Dose: 2.5 Millicuries
- Inj. date (date/month - year): 13/06 - 2007
- Inj. time: 11:18:00
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 10 PtMinute

4.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

4.3 Recon

4.3.1 Recon 1

- Series description: PET Brain
- Recon range (bed): ups..
- Output image type: Corrected
- Recon method: Iterative
- Iterations: 6
- Subsets: 16
- Image size: 336
- Zoom: 2
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off