

# 3\_PET\_AC\_LM\_FET\_Brain\_4.Adult\_PET4

30. maj 2013

# Indhold

|          |                     |          |
|----------|---------------------|----------|
| <b>1</b> | <b>Topogram</b>     | <b>1</b> |
| 1.1      | Routine . . . . .   | 1        |
| 1.2      | Scan . . . . .      | 1        |
| <b>2</b> | <b>Lavdosis CT</b>  | <b>1</b> |
| 2.1      | Routine . . . . .   | 1        |
| 2.2      | Scan . . . . .      | 1        |
| 2.3      | Recons . . . . .    | 2        |
| 2.3.1    | Recon 1 . . . . .   | 2        |
| 2.3.2    | Recon 2 . . . . .   | 2        |
| <b>3</b> | <b>Pause</b>        | <b>2</b> |
| <b>4</b> | <b>PET Brain LM</b> | <b>2</b> |
| 4.1      | Routine . . . . .   | 2        |
| 4.2      | Scan . . . . .      | 3        |
| 4.3      | Recons . . . . .    | 3        |
| 4.3.1    | Recon 1 . . . . .   | 3        |
| 4.3.2    | Recon 2 . . . . .   | 3        |

# **1 Topogram**

## **1.1 Routine**

- mA: 35
- kV: 120
- Topogram length: 256 mm
- Tube position: Lateral

## **1.2 Scan**

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 256 mm
- Direction: Craniocaudal
- Tube position: Lateral
- API: API
- Kernel: 20
- Window: Topogram

# **2 Lavdosis CT**

## **2.1 Routine**

- Eff. mAs: 350
- kV: 120
- CARE Dose4D: On
- CareDoseType: CareDoseAEC
- CTDIvol: 59.7805mGy
- Scan time: 26.000 s
- Delay: 2.000 s
- Slice: 3 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

## **2.2 Scan**

- Quality ref. mAs: 350
- Eff. mAs: 350
- kV: 120
- Scan time: 26.000 s
- Rotation time: 1.000 s
- Delay: 2.000 s
- Slice: 3 mm
- Pitch: 1.50

- Direction: Caudocranial

## **2.3 Recons**

### **2.3.1 Recon 1**

- Series description: Lavdosis CT RTD
- Slice: 1.5
- Kernel: H19s
- Window: Cerebrum
- Extended FoV: Off
- FoV: 300
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 1.500
- No. of images: 148

### **2.3.2 Recon 2**

- Series description: AC Lavdosis CT
- Slice: 3
- Kernel: H19s
- Window: Cerebrum
- Extended FoV: Off
- FoV: 300
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 3.000
- No. of images: 74

## **3 Pause**

## **4 PET Brain LM**

### **4.1 Routine**

- Isotope: F-18

- Pharm.: FET
- Inj. Dose: 185000000 Bequerels
- Scan mode: List mode
- Scan range: Do not match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 20.000000 Minutes

## **4.2 Scan**

- Autoload: Off
- Rebinner LUT: Off
- Scan output: List mode
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): n/a
- ULD (keV): n/a

## **4.3 Recons**

### **4.3.1 Recon 1**

- Series description: FET stat 20-40 min 5 mm
- Recon range (bed): 1 to 1
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

### **4.3.2 Recon 2**

- Series description: FET stat 25-35 min 5 mm
- Recon range (bed): 1 to 1
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12

- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off