# PETCT\_AC\_Brain\_Ped.MlChild\_PET5

20. februar 2013

# Indhold

1	Topogram 1				
	1.1	Routine	1		
	1.2	Scan	1		
2	CT Brain				
	2.1	Routine	1		
	2.2	Scan	2		
	2.3	Recons			
		2.3.1 Recon 1	2		
3	Pau	se :	3		
4	PE	Brain	3		
	4.1	Routine	3		
	4.2	Scan	3		
	4.3	Recons	4		
		4.3.1 Recon 1	4		

# 1 Topogram

# 1.1 Routine

• mA: 50

• kV: 80

• Topogram length: 256 mm

• Tube position: Lateral

#### 1.2 Scan

• mA: 50

• kV: 80

• Delay: 4s

• Topogram length: 256 mm

• Direction: Craniocaudal

• Tube position: Lateral

• API: None

• Kernel: 20s standar

• Window: Topogram Head

# 2 CT Brain

## 2.1 Routine

• Eff. mAs: 85

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

• CTDlvol: 14.399mGy

• Scan time: 17.000 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

# 2.2 Scan

• Quality ref. mAs: 380

• Eff. mAs: 85

• kV: 120

• Scan time: 17.000 s

• Rotation time: 1.000 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 1.2000000

• Direction: Caudocranial

## 2.3 Recons

#### 2.3.1 Recon 1

• Series description: AC CT

• Slice: 3.00

• Kernel: H19s PET very smooth

• Window: Cerebrum

• Extended FoV: Off

• FoV: 300

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Caudocranial

• Recon increment: 1.500

• No. of images: 148

# 3 Pause

## 4 PET Brain

#### 4.1 Routine

• Isotope: F-18

• Pharm.: FDG

• Inj. Dose: 2.5 Millicuries

• Scan mode: Sinogram

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 10 Minutes

## 4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

#### 4.3 Recons

#### 4.3.1 Recon 1

- Series description: PET Brain
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: Iterative
- Iterations: 6
- Subsets: 16
- Image size: 336
- Zoom: 2
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off