# $PETCT\_WB\_FDG\_IV\_15\_25\_kg\_Child.MlChild\_PET5$

7. marts 2013

## Indhold

1	Topogram					
	1.1	Routine	1			
	1.2	Scan	1			
2	CT		1			
	2.1	Routine	1			
	2.2	Scan	1			
	2.3	Recons	2			
		2.3.1 Recon 1	2			
		2.3.2 Recon 2	2			
3	Pau	se	2			
4	PET	$^{\circ}$ WB	2			
	4.1	Routine	2			
	4.2	Scan	3			
	4.3	Recons	3			
		4.3.1 Recon 1	3			
		4.3.2 Recon 2	3			

## 1 Topogram

#### 1.1 Routine

mA: 35kV: 120

 $\bullet \;$  Topogram length: 0 mm

• Tube position: Top

#### 1.2 Scan

mA: 35kV: 120Delay: 4s

Topogram length: 0 mmDirection: Craniocaudal

• Tube position: Top

• API: None

• Kernel: 20s standar

• Window: Topogram Body

## 2 CT

#### 2.1 Routine

• Eff. mAs: 35

• kV: 120

• CARE Dose4D: On

• CareDoseType: CareDoseAEC

CTDlvol: 2.35627mGyScan time: 10.510 sDelay: 20.000 s

Delay: 20.000 sSlice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

#### 2.2 Scan

• Quality ref. mAs: 35

• Eff. mAs: 35

• kV: 120

Scan time: 10.510 sRotation time: 0.500 s

Delay: 20.000 sSlice: 3.00 mmPitch: 1.2000000

• Direction: Craniocaudal

#### 2.3 Recons

#### 2.3.1 Recon 1

- Series description: AC CT
- Slice: 3.00
- Kernel: B19f PET very smooth
- Window: Baby Abdomen
- Extended FoV: On
- FoV: 700
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 224

#### 2.3.2 Recon 2

- Series description: CT WB 3.0 B30f
- Slice: 3.00
- Kernel: B30f medium smooth
- Window: Baby Abdomen
- Extended FoV: Off
- FoV: 300
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 224

## 3 Pause

## 4 PET WB

## 4.1 Routine

• Isotope: F-18

- Pharm.: FDG
- Inj. Dose: 1 MegaBequerels
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: 4
- Scan duration/bed: 3 Minutes

#### 4.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 425
- ULD (keV): 650

#### 4.3 Recons

#### 4.3.1 Recon 1

- Series description: PET WB Image check
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

#### 4.3.2 Recon 2

- Series description: PET WB Corrected True X 2 mm
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 2
- Subsets: 21

• Image size: 336

• Zoom: 1

Filter: GaussianFWHM (mm): 2Offset X: 0 mmOffset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: On

Match CT slice location: On Save intermediate data: Off