3_PET_AC_FDG_Brain_Neurostat.Adult_PET4

11. januar 2013

Indhold

1 Topogram

1.1 Routine

- mA: 35

- kV: 120

- Topogram length: 256 mm

- Tube position: Lateral

1.2 Scan

- mA: 35

- kV: 120

- Delay: 4s

- Topogram length: 256 mm

- Direction: Craniocaudal

- Tube position: Lateral

- API: API

- Kernel: 20

- Window: Topogram

2 CT Brain

2.1 Routine

- Eff. mAs: 30

- kV: 120

- CARE Dose4D: On

- CareDoseType: CareDoseAEC

- CTDlvol: 5.124mGy

- Scan time: 2.640 s

- Delay: 2.000 s

- Slice: 3 mm

- No. of images: Samme som i foerste recon, slet?(y/n)

- Tilt: 0.0 grader

2.2 Scan

- Quality ref. mAs: 30

- Eff. mAs: 30

- kV: 120

- Scan time: 2.640 s

- Rotation time: 1.000 s

- Delay: 2.000 s

- Slice: 3 mm

- Pitch: 1.50

- Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

- Series description: tmp CT Brain RTD

- Slice: 5

- Kernel: H19s

- Window: Cerebrum

- Extended FoV: Off

- FoV: 300

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Craniocaudal

- Recon increment: 5.000

- No. of images: 1

2.3.2 Recon 2

- Series description: AC CT

- Slice: 3

- Kernel: H19s

- Window: Cerebrum

- Extended FoV: Off

- FoV: 300

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Craniocaudal

- Recon increment: 3.000

- No. of images: 1

3 Pause

4 PET Brain

4.1 Routine

- Isotope: F-18

- Pharm.: Fluorodeoxyglucose

- Inj. Dose: 200000000 Bequerels

- Inj. date (date/month - year): 15/05 - 2012

- Inj. time: 09:15:23

- Scan mode: Sinogram

- Scan range:

- No. of beds: Not given. Check recon range

- Scan duration/bed: 10.000000 Minutes

4.2 Scan

- Autoload: On

- Rebinner LUT: Off

- Scan output: Sinogram

- Sinogram mode: false

- Input trigger signal: None

- LLD (keV): n/a

- ULD (keV): n/a

4.3 Recons

4.3.1 Recon 1

- Series description: PET Brain Corrected 3 mm zoom 3

- Recon range (bed): Automatic/All/dunno

- Output image type: Corrected

- Recon method: Iterative3D

- Iterations: 4

- Subsets: 24

- Image size: 400

- Zoom: 3.000000

- Filter: Gaussian

- FWHM (mm): 3.000000

- Offset X: 0 mm

- Offset Y: 0 mm

- Attenuation correction: Off ()

- Scatter correction: On

- Match CT slice location: On

- Save intermediate data: Off

4.3.2 Recon 2

- Series description: Neurostat

- Recon range (bed): Automatic/All/dunno

- Output image type: Corrected

- Recon method: Iterative3D

- Iterations: 4

- Subsets: 8

- Image size: 128

- Zoom: 3.000000

- Filter: Gaussian

- FWHM (mm): 8.000000

- Offset X: 0 mm

- Offset Y: 0 mm

- Attenuation correction: Off ()

- Scatter correction: On

- Match CT slice location: On

- Save intermediate data: Off