$Hjerne_FET_PET_Dyn_LM_6.MlAdult_PET6$

27. februar 2013

Indhold

1	Topogram				
	1.1	Routine	1		
	1.2	Scan	1		
2	AC CT Brain				
	2.1	Routine	1		
	2.2	Scan	2		
	2.3	Recons	2		
		2.3.1 Recon 1	2		
		2.3.2 Recon 2	3		
3	Pau	${f se}$	4		
4	PET	Γ Brain	4		
	4.1	Routine	4		
	4.2	Scan	4		
	4.3	Recons	4		
		4.3.1 Recon 1	4		

1 Topogram

1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 256 mm

• Tube position: Lateral

1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 256 mm

• Direction: Craniocaudal

• Tube position: Lateral

• API: None

• Kernel: 20s standar

• Window: Topogram Head

2 AC CT Brain

2.1 Routine

• Eff. mAs: 30

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

• CTDlvol: 4.71218mGy

• Scan time: 11.230 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

2.2 Scan

• Quality ref. mAs: 380

• Eff. mAs: 30

• kV: 120

• Scan time: 11.230 s

• Rotation time: 1.000 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 1.2000000

• Direction: Caudocranial

2.3 Recons

2.3.1 Recon 1

• Series description: AC CT

• Slice: 3.00

• Kernel: H19s PET very smooth

• Window: Cerebrum

• Extended FoV: Off

• FoV: 300

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Caudocranial

• Recon increment: 3.000

• No. of images: 74

2.3.2 Recon 2

• Series description: CT 8mm

• Slice: 3.00

• Kernel: H19s PET very smooth

• Window: Cerebrum

• Extended FoV: Off

• FoV: 300

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Caudocranial

• Recon increment: 8.000

• No. of images: 28

3 Pause

4 PET Brain

4.1 Routine

• Isotope: F-18

• Pharm.: FET

• Inj. Dose: 224 MegaBequerels

• Scan mode: List mode

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 40 Minutes

4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: List mode

• Sinogram mode: Trues

• Input trigger signal: None

• LLD (keV): 435

• ULD (keV): 650

4.3 Recons

4.3.1 Recon 1

• Series description: PET FET Brain corrected

• Recon range (bed): -1 to -1

• Output image type: NoRecon

• Recon method: Iterative3D

• Iterations: 6

• Subsets: 16

• Image size: 336

• Zoom: 3

• Filter: Gaussian

• FWHM (mm): 3

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: On

• Match CT slice location: On

• Save intermediate data: Off