

4_PETCT_Brain.Adult

8. oktober 2012

Indhold

1	Topogram	1
1.1	Routine	1
1.2	Scan	1
2	CT WB	1
2.1	Routine	1
2.2	Scan	2
2.3	Recon	2
2.3.1	Recon 1	2
2.3.2	Recon 2	3
2.3.3	Recon 3	4
3	Pause	4
4	PET WB	4
4.1	Routine	4
4.2	Scan	5
4.3	Recon	5
4.3.1	Recon 1	5

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 256 mm
- Tube position: Lateral

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 256 mm
- Direction: Craniocaudal
- Tube position: Lateral
- API:
- Kernel: T20s
- Window: Topogra

2 CT WB

2.1 Routine

- Eff. mAs: 340
- kV: 120
- CARE Dose4D: Off
- CTDIvol: 58.072mGy
- Scan time: 68.730 s

- Delay: 2.000 s
- Slice: 3 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

2.2 Scan

- Quality ref. mAs: 350
- Eff. mAs: 340
- kV: 120
- Scan time: 68.730 s
- Rotation time: 1.000 s
- Delay: 2.000 s
- Slice: 3 mm
- Pitch: 0.55
- Direction: Craniocaudal

2.3 Recon

2.3.1 Recon 1

- Series description: tmp CT Brain RTD
- Slice: 5
- Kernel: "H19s"
- Window: "Cerebrum"
- Extended FoV: false
- FoV: 300
- Center X: 0
- Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 5.000
- No. of images: 45

2.3.2 Recon 2

- Series description: AC CT Brain
- Slice: 3
- Kernel: "H19s"
- Window: "Cerebrum"
- Extended FoV: false
- FoV: 300
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 1.500
- No. of images: 148

2.3.3 Recon 3

- Series description: CT Brain 3.0 H31s
- Slice: 3
- Kernel: "H31s"
- Window: "Cerebrum"
- Extended FoV: false
- FoV: 250
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 1.500
- No. of images: 148

3 Pause

4 PET WB

4.1 Routine

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- Inj. Dose: 1
- Inj. date (date/month - year): 02/10 - 2008
- Inj. time: 06:23:15

- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: N/A
- Scan duration/bed: 10.000000 Minutes

4.2 Scan

- Autoload: true
- Rebinner LUT: N/A
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- ULD (keV): N/A

4.3 Recon

4.3.1 Recon 1

- Series description: PET Brain
- Recon range (bed): ups..
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 24
- Image size: 400
- Zoom: 2.000000
- Filter: Gaussian
- FWHM (mm): 4.000000

- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false