# $HUND\_64Cu\_Liposom\_Dag2.MlAdult$

1. oktober 2012

## Indhold

1	Top	ogram	1		
			1		
	1.2	Scan	1		
<b>2</b>	CT WB				
	2.1	Routine	1		
	2.2	Scan	2		
	2.3	Recon	2		
		2.3.1 Recon 1	2		
		2.3.2 Recon 2	3		
3	Pau	${f se}$	4		
4	PET	$\Gamma  ext{ WB}$	4		
	4.1	Routine	4		
	4.2	Scan	4		
	4.3	Recon	1		
		4.3.1 Recon 1	5		

## 1 Topogram

#### 1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 0 mm
- Tube position: Top

#### 1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 0 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: T80s sharp
- Window: Topogram Body

## 2 CT WB

## 2.1 Routine

- Eff. mAs: 170
- kV: 12
- CARE Dose4D: Off
- CTDlvol: 11.4448mGy
- $\bullet$  Scan time: 1.2000000 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 170 grader

## 2.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 170

• kV: 120

• Scan time: 12.680 s

• Rotation time: 0.500 s

• Delay: 4.000 s

• Slice: 0.6 mm

• Pitch: 1.2000000

• Direction: Craniocaudal

#### 2.3 Recon

#### 2.3.1 Recon 1

 $\bullet\,$  Series description: AC CT

• Slice: 3.00

• Kernel: "B19f PET very smooth"

• Window: "Abdomen"

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 274

#### 2.3.2 Recon 2

• Series description: CT 3.0 B30f medium smooth

• Slice: 3.00

• Kernel: "B30f medium smooth"

• Window: "Abdomen"

• Extended FoV: Off

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 274

## 3 Pause

## 4 PET WB

## 4.1 Routine

• Isotope: Cu-64

• Pharm.: liposome

• Inj. Dose: 1 MegaBequerels

 $\bullet\,$  Inj. date (date/month - year): 08/02 - 2012

• Inj. time: 16:04:00

• Scan mode: Sinogram

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 3 PtMinute

#### 4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

• Input trigger signal: None

• LLD (keV): 435

• ULD (keV): 650

## 4.3 Recon

#### 4.3.1 Recon 1

- Series description: PET HEART True X Day2
- Recon range (bed): ups..
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 3
- Subsets: 21
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off