# 3\_PET\_AC\_LM\_FET\_Brain\_HGREB.Adult\_PET4

23. januar 2013

# Indhold

1	Top	ogram		
	1.1	Routine		
	1.2	Scan		
2	Lav	dosis CT		
	2.1	Routine		
	2.2	Scan		
	2.3	Recons		
		2.3.1 Recon 1		
		2.3.2 Recon 2		
3	Pau	se		
4	PET Brain LM			
	4.1	Routine		
	4.2	Scan		
	4.3	Recons		
		4.3.1 Recon 1		
		4.3.2 Recon 2		

## 1 Topogram

## 1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 256 mm

• Tube position: Lateral

#### 1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 256 mm

• Direction: Craniocaudal

• Tube position: Lateral

• API: API

• Kernel: 20

• Window: Topogram

## 2 Lavdosis CT

## 2.1 Routine

• Eff. mAs: 30

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

• CTDlvol: 5.124mGy

• Scan time: 26.200 s

• Delay: 2.000 s

• Slice: 3 mm

 $\bullet$  No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

## 2.2 Scan

• Quality ref. mAs: 350

• Eff. mAs: 30

• kV: 120

 $\bullet$  Scan time: 26.200 s

• Rotation time: 1.000 s

• Delay: 2.000 s

• Slice: 3 mm

• Pitch: 1.50

• Direction: Caudocranial

## 2.3 Recons

#### 2.3.1 Recon 1

• Series description: tmp Lavdosis CT RTD

 $\bullet$  Slice: 5

• Kernel: H19s

• Window: Cerebrum

• Extended FoV: Off

• FoV: 300

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Caudocranial

• Recon increment: 5.000

• No. of images: 45

#### 2.3.2 Recon 2

• Series description: AC Lavdosis CT

• Slice: 3

• Kernel: H19s

• Window: Cerebrum

• Extended FoV: Off

• FoV: 300

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Caudocranial

• Recon increment: 3.000

• No. of images: 74

## 3 Pause

## 4 PET Brain LM

## 4.1 Routine

• Isotope: F-18

• Pharm.: FET

• Inj. Dose: 237000000 Bequerels

• Scan mode: List mode

• Scan range: Do not match CT FOV

• No. of beds: Not given. Check recon range

• Scan duration/bed: 20.000000 Minutes

## 4.2 Scan

• Autoload: Off

• Rebinner LUT: Off

• Scan output: List mode

• Sinogram mode: false

• Input trigger signal: None

• LLD (keV): n/a

• ULD (keV): n/a

#### 4.3 Recons

#### 4.3.1 Recon 1

• Series description: FET stat 1 mm

• Recon range (bed): 1 to 1

• Output image type: Corrected

- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 1.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

#### 4.3.2 Recon 2

- Series description: FET stat 5 mm4
- Recon range (bed): 1 to 1
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400
- Zoom: 2.500000
- Filter: Gaussian
- FWHM (mm): 5.000000
- Offset X: 0 mm

• Offset Y: 0 mm

 $\bullet$  Attenuation correction: Off ()

• Scatter correction: On

• Match CT slice location: On

• Save intermediate data: Off