PROJ_COLON_FLT_PET_5.MlAdult

23. oktober 2012

Indhold

1	1 Topogram	1
	1.1 Routine	 1
	1.2 Scan	 1
2	2 Lavdos CT WB	1
	2.1 Routine	 1
	2.2 Scan	 2
	2.3 Recons	 2
	2.3.1 Recon 1	 2
3	3 Pause	3
4	4 PET WB	3
	4.1 Routine	 3
	4.2 Scan	 4
	4.3 Recons	 4
	4.3.1 Recon 1	 4
	4.3.2 Recon 2	 5

1 Topogram

1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 1024 mm

• Tube position: Top

1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 270 mm

• Direction: Craniocaudal

• Tube position: Top

• API: None

• Kernel: 80s shar

• Window: Topogram Body

2 Lavdos CT WB

2.1 Routine

• Eff. mAs: 40

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

• CTDlvol: 2.70069mGy

• Scan time: 18.670 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 40 grader

2.2 Scan

• Quality ref. mAs: 50

• Eff. mAs: 40

• kV: 120

 \bullet Scan time: 18.670 s

• Rotation time: 0.500 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 0.8000000

• Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

• Series description: AC CT

• Slice: 3.00

• Kernel: B19f PET very smooth

• Window: Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 274

3 Pause

4 PET WB

4.1 Routine

• Isotope: F-18

• Pharm.: FLT

• Inj. Dose: 1 MegaBequerels

• Inj. date (date/month - year): 18/01 - 2012

• Inj. time: 09:22:00

• Scan mode: Sinogram

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 150 Seconds

4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

• Input trigger signal: None

• LLD (keV): 435

• ULD (keV): 650

4.3 Recons

4.3.1 Recon 1

• Series description: PET WB Image check

• Recon range (bed): Automatic/All/dunno

• Output image type: Corrected

• Recon method: Iterative

• Iterations: 4

• Subsets: 8

• Image size: 168

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 4

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: On

• Match CT slice location: On

• Save intermediate data: Off

4.3.2 Recon 2

- \bullet Series description: PET FLT Corr. True X 2 mm
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 3
- Subsets: 21
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off