$TUMOR_HUND_FDG.MlAdult_PET5$

11. januar 2013

Indhold

| 1 | Topogram 1 | | | |
|----------|-------------------|------------------|---|--|
| | 1.1 | Routine | 1 | |
| | 1.2 | Scan | 1 | |
| 2 | CT | WB | 1 | |
| | 2.1 | Routine | 1 | |
| | 2.2 | Scan | 2 | |
| | 2.3 | Recons | 2 | |
| | | 2.3.1 Recon 1 | 2 | |
| | | 2.3.2 Recon 2 | 3 | |
| 3 | Pau | se | 4 | |
| 4 | \mathbf{PE}^{r} | T TUMOR | 4 | |
| _ | 4.1 | | 4 | |
| | 4.2 | | 4 | |
| | 4.3 | | 5 | |
| | | | 5 | |
| | | 4.3.2 Recon 2 | 5 | |
| 5 | Pau | ${f se}$ | 6 | |
| 6 | \mathbf{PE}^{r} | Γ WB | 6 | |
| Ü | 6.1 | | 6 | |
| | 6.2 | | 7 | |
| | 6.3 | | 7 | |
| | 0.0 | | 7 | |
| | | | 8 | |
| 7 | Pau | ${f se}$ | 8 | |
| 8 | Dvr | \mathbf{Multi} | 8 | |
| - | 8.1 | | 8 | |
| | 8.2 | | 9 | |
| | 8.3 | Recons | 9 | |
| | | 8 3 1 Recon 1 | O | |

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 0 mm
- Tube position: Top

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 0 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: 80s shar
- Window: Topogram Body

2 CT WB

2.1 Routine

- Eff. mAs: 170
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDlvol: 11.4448mGy

• Scan time: 12.680 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

2.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 170

• kV: 120

 \bullet Scan time: 12.680 s

• Rotation time: 0.500 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 1.2000000

• Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

• Series description: AC CT

• Slice: 3.00

• Kernel: B19f PET very smooth

• Window: Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 274

2.3.2 Recon 2

• Series description: CT WB 3.0 B30f

• Slice: 3.00

• Kernel: B30f medium smooth

• Window: Abdomen

• Extended FoV: Off

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 274

3 Pause

4 PET TUMOR

4.1 Routine

• Isotope: F-18

• Pharm.: FDG

• Inj. Dose: 200 MegaBequerels

 \bullet Inj. date (date/month - year): 20/06 - 2011

• Inj. time: 17:13:00

• Scan mode: Sinogram

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 10 Minutes

4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

• Input trigger signal: None

• LLD (keV): 435

• ULD (keV): 650

4.3 Recons

4.3.1 Recon 1

- Series description: PET TUMOR Image check
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

4.3.2 Recon 2

- Series description: PET TUMOR True X 2.0 mm
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 3
- Subsets: 21

- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: Off
- Match CT slice location: On
- Save intermediate data: Off

5 Pause

6 PET WB

6.1 Routine

- Isotope: F-18
- Pharm.: FDG
- Inj. Dose: 200 MegaBequerels
- Inj. date (date/month year): 20/06 2011
- Inj. time: 17:13:00
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 2 Minutes

6.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

• Input trigger signal: None

• LLD (keV): 435

• ULD (keV): 650

6.3 Recons

6.3.1 Recon 1

• Series description: PET WB Image check

• Recon range (bed): Automatic/All/dunno

• Output image type: Corrected

• Recon method: Iterative

• Iterations: 4

• Subsets: 8

• Image size: 128

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 4

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: On

• Match CT slice location: On

• Save intermediate data: Off

6.3.2 Recon 2

- Series description: PET WB True X 2.0 mm
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 3
- Subsets: 21
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: Off
- Match CT slice location: On
- Save intermediate data: Off

7 Pause

8 DynMulti

8.1 Routine

- Eff. mAs: 170
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDose

 \bullet CTDlvol: 1032.37mGy

• Scan time: 90.100 s

• Delay: 4.000 s

• Slice: 7.20 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

8.2 Scan

• Quality ref. mAs: 100

• Eff. mAs: 170

• kV: 120

• Scan time: 90.100 s

• Rotation time: 1.000 s

• Delay: 4.000 s

• Slice: 7.20 mm

• Pitch: 1.0000000

• Direction: Craniocaudal

8.3 Recons

8.3.1 Recon 1

• Series description: DynMulti 7.2 B30s

• Slice: 7.20

• Kernel: B30s medium smooth

• Window: Abdomen

• Extended FoV: Off

• FoV: 0

• Center X: -3

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 1.000

• No. of images: 360