4_PET_Terapi_Boost_IVAdult.Adult_PET4

7. marts 2013

Indhold

1 Topogram **1**
- 1.1 Routine . . . 1
- 1.2 Scan . . . 1

2 LavdoseCTWB **1**
- 2.1 Routine . . . 1
- 2.2 Scan . . . 1
- 2.3 Recons . . . 2
  - 2.3.1 Recon 1 . . . 2
  - 2.3.2 Recon 2 . . . 2

3 Pause **2**

4 PET WB **2**
- 4.1 Routine . . . 2
- 4.2 Scan . . . 3
- 4.3 Recons . . . 3
  - 4.3.1 Recon 1 . . . 3
  - 4.3.2 Recon 2 . . . 3

5 Pause **4**

6 PreMonitoring **4**
- 6.1 Routine . . . 4
- 6.2 Scan . . . 4
- 6.3 Recons . . . 4
  - 6.3.1 Recon 1 . . . 4

7 Monitoring **5**
- 7.1 Routine . . . 5
- 7.2 Scan . . . 5
- 7.3 Recons . . . 5
  - 7.3.1 Recon 1 . . . 5

8 WB CT **6**
- 8.1 Routine . . . 6
- 8.2 Scan . . . 6
- 8.3 Recons . . . 7
  - 8.3.1 Recon 1 . . . 7
  - 8.3.2 Recon 2 . . . 7
  - 8.3.3 Recon 3 . . . 7

8.3.4 Recon 4 8
8.3.5 Recon 5 8

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 1024 mm
- Tube position: Top

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 1024 mm
- Direction: Craniocaudal
- Tube position: Top
- API: API
- Kernel: 80
- Window: Topogram

2 LavdoseCTWB

2.1 Routine

- Eff. mAs: 50
- kV: 120
- CARE Dose4D: On
- CareDoseType: CareDoseAEC
- CTDIvol: 3.65446mGy
- Scan time: 32.745 s
- Delay: 2.000 s
- Slice: 2 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

2.2 Scan

- Quality ref. mAs: 50
- Eff. mAs: 50
- kV: 120
- Scan time: 32.745 s
- Rotation time: 0.500 s
- Delay: 2.000 s
- Slice: 2 mm
- Pitch: 0.80

- Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

- Series description: LavdoseCTWB RTD
- Slice: 4
- Kernel: B19f
- Window: Abdomen
- Extended FoV: Off
- FoV: 500
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 4.000
- No. of images: 245

2.3.2 Recon 2

- Series description: AC lavCT 2 mm
- Slice: 2
- Kernel: B19f
- Window: Abdomen
- Extended FoV: On
- FoV: 780
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 489

3 Pause

4 PET WB

4.1 Routine

- Isotope: F-18

- Pharm.: Fluorodeoxyglucose
- Inj. Dose: 0 Bequerels
- Scan mode: Sinogram
- Scan range: Do not match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 2.000000 Minutes

4.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): n/a
- ULD (keV): n/a

4.3 Recons

4.3.1 Recon 1

- Series description: PET WB Image check HD
- Recon range (bed): 1 to 7
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 8
- Image size: 200
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

4.3.2 Recon 2

- Series description: PET WB Corr TrueX+TOF 2 mm
- Recon range (bed): 1 to 7
- Output image type: Corrected
- Recon method: TrueXTOF
- Iterations: 2
- Subsets: 21

- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

5 Pause

6 PreMonitoring

6.1 Routine

- Eff. mAs: 20
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDlvol: 1.18627mGy
- Scan time: 0.500 s
- Delay: 2.000 s
- Slice: 10 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

6.2 Scan

- Quality ref. mAs: 225
- Eff. mAs: 20
- kV: 120
- Scan time: 0.500 s
- Rotation time: 0.500 s
- Delay: 2.000 s
- Slice: 10 mm
- Pitch: 0.00
- Direction: Craniocaudal

6.3 Recons

6.3.1 Recon 1

- Series description: PreMonitoring 10.0 eFoV
- Slice: 10

- Kernel: B19s
- Window: Abdomen
- Extended FoV: On
- FoV: 780
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 10.000
- No. of images: 1

7 Monitoring

7.1 Routine

- Eff. mAs: 20
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDlvol: 35.5881mGy
- Scan time: 0.500 s
- Delay: 10.000 s
- Slice: 10 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

7.2 Scan

- Quality ref. mAs: 225
- Eff. mAs: 20
- kV: 120
- Scan time: 0.500 s
- Rotation time: 0.500 s
- Delay: 10.000 s
- Slice: 10 mm
- Pitch: 0.00
- Direction: Craniocaudal

7.3 Recons

7.3.1 Recon 1

- Series description: Monitoring 10.0 eFoV

- Slice: 10
- Kernel: B19s
- Window: Abdomen
- Extended FoV: On
- FoV: 780
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 10.000
- No. of images: 30

8 WB CT

8.1 Routine

- Eff. mAs: 200
- kV: 120
- CARE Dose4D: On
- CareDoseType: CareDoseAEC
- CTDlvol: 14.6179mGy
- Scan time: 32.745 s
- Delay: 10.000 s
- Slice: 2 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

8.2 Scan

- Quality ref. mAs: 200
- Eff. mAs: 200
- kV: 120
- Scan time: 32.745 s
- Rotation time: 0.500 s
- Delay: 10.000 s
- Slice: 2 mm
- Pitch: 0.80
- Direction: Craniocaudal

8.3 Recons

8.3.1 Recon 1

- Series description: WB CT RTD
- Slice: 4
- Kernel: B19f
- Window: Abdomen
- Extended FoV: Off
- FoV: 500
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 4.000
- No. of images: 245

8.3.2 Recon 2

- Series description: AC CT 2 mm
- Slice: 2
- Kernel: B19f
- Window: Abdomen
- Extended FoV: On
- FoV: 780
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 489

8.3.3 Recon 3

- Series description: CT WB 2.0 B30f
- Slice: 2
- Kernel: B30f
- Window: Abdomen
- Extended FoV: Off
- FoV: 500

- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 489

8.3.4 Recon 4

- Series description: CT WB 2.0 cor B30f
- Slice: 2
- Kernel: B30f
- Window: Abdomen
- Extended FoV: Off
- FoV: 977
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: 3D
- Recon Axis: Coronal
- Image order: AP
- Recon increment: 2.000
- No. of images: 251

8.3.5 Recon 5

- Series description: CT WB 2.0 sag B30f
- Slice: 2
- Kernel: B30f
- Window: Abdomen
- Extended FoV: Off
- FoV: 977
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: 3D
- Recon Axis: Sagittal
- Image order: LeftToRight
- Recon increment: 2.000
- No. of images: 489