# $Hjerne\_Dotatoc\_PET\_5.MlAdult$

27. september 2012

# Indhold

| 1 |               | ogram         |
|---|---------------|---------------|
|   | 1.1           | Routine       |
|   | 1.2           | Scan          |
| 2 | $\mathbf{CT}$ |               |
|   | 2.1           | Routine       |
|   | 2.2           | Scan          |
|   | 2.3           | Recon         |
|   |               | 2.3.1 Recon 1 |
| 3 | Pau           | ${f se}$      |
| 4 | PE:           | r WB          |
|   | 4.1           | Routine       |
|   | 4.2           | Scan          |
|   | 4.3           | Recon         |
|   |               | 4.3.1 Recon 1 |

### 1 Topogram

#### 1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 256 mm

• Tube position: Lateral

#### 1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 256 mm

• Direction: Craniocaudal

• Tube position: Lateral

• API: None

• Kernel: T20s standard

• Window: Topogram Head

## 2 CT WB

#### 2.1 Routine

• Eff. mAs: 30

• kV: 8

• CARE Dose4D: Off

• CTDlvol: 4.12323mGy

 $\bullet$  Scan time: 1.2000000 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 30 grader

#### 2.2 Scan

• Quality ref. mAs: 380

• Eff. mAs: 30

• kV: 120

• Scan time: 8.010 s

• Rotation time: 1.000 s

• Delay: 4.000 s

• Slice: 0.6 mm

• Pitch: 1.2000000

• Direction: Craniocaudal

#### 2.3 Recon

#### 2.3.1 Recon 1

 $\bullet\,$  Series description: AC CT

• Slice: 3.00

• Kernel: "H19s PET very smooth"

• Window: "Cerebrum"

• Extended FoV: Off

• FoV: 300

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Caudocranial

• Recon increment: 3.000

• No. of images: 74

#### 3 Pause

#### 4 PET WB

#### 4.1 Routine

• Isotope: Ga-68

• Pharm.: Dotatoc

• Inj. Dose: 1 MegaBequerels

• Inj. date (date/month - year): 15/05 - 2012

• Inj. time: 09:34:00

• Scan mode: List mode

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 15 PtMinute

#### 4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: List mode

- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

#### 4.3 Recon

#### 4.3.1 Recon 1

- Series description: PET Brain corrected 3 mm zoom 3
- Recon range (bed): ups..
- Output image type: NoRecon
- Recon method: Iterative3D
- Iterations: 6
- Subsets: 16
- Image size: 336
- Zoom: 3
- Filter: Gaussian
- FWHM (mm): 3
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off