# PROJ\_NET\_FLT\_PET\_CT\_IV\_6.MlAdult\_PET6

20. februar 2013

# Indhold

1	Topogram 1				
	1.1	Routine	1		
	1.2	Scan	1		
<b>2</b>	$\mathbf{CT}$	$_{ m WB}$	1		
	2.1	Routine	1		
	2.2	Scan	2		
	2.3	Recons	2		
		2.3.1 Recon 1	2		
		2.3.2 Recon 2	3		
		2.3.3 Recon 3	4		
		2.3.4 Recon 4	4		
3	Pau	se	5		
4	PE	$^{\circ}$ WB	5		
	4.1	Routine	5		
	4.2	Scan	5		
	4.3	Recons	6		
		4.3.1 Recon 1	6		
		4.3.2 Recon 2	7		

## 1 Topogram

## 1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 0 mm

• Tube position: Top

#### 1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 0 mm

• Direction: Craniocaudal

• Tube position: Top

• API: None

• Kernel: 80s shar

• Window: Topogram Body

## 2 CT WB

## 2.1 Routine

• Eff. mAs: 225

• kV: 120

• CARE Dose4D: On

• CareDoseType: CareDoseAEC

• CTDlvol: 15.1475mGy

• Scan time: 12.680 s

• Delay: 75.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

## 2.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 225

• kV: 120

 $\bullet$  Scan time: 12.680 s

• Rotation time: 0.500 s

• Delay: 75.000 s

• Slice: 3.00 mm

• Pitch: 1.2000000

• Direction: Craniocaudal

## 2.3 Recons

#### 2.3.1 Recon 1

• Series description: AC CT

• Slice: 3.00

• Kernel: B19f PET very smooth

• Window: Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 274

#### 2.3.2 Recon 2

• Series description: CT WB 3.0 B30f

• Slice: 3.00

• Kernel: B30f medium smooth

• Window: Abdomen

• Extended FoV: Off

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 274

#### 2.3.3 Recon 3

 $\bullet$  Series description: CT WB 3.0 cor

• Slice: 3.00

• Kernel: B30f medium smooth

• Window: Abdomen

• Extended FoV: Off

• FoV: 821

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: 3D

• Recon Axis: Coronal

• Image order: AP

• Recon increment: 3.000

• No. of images: 167

#### 2.3.4 Recon 4

• Series description: CT WB 3.0 sag

• Slice: 3.00

• Kernel: B30f medium smooth

• Window: Abdomen

• Extended FoV: Off

• FoV: 821

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: 3D

• Recon Axis: Sagittal

• Image order: LeftToRight

• Recon increment: 3.000

• No. of images: 274

## 3 Pause

## 4 PET WB

## 4.1 Routine

• Isotope: F-18

• Pharm.: FLT

• Inj. Dose: 1 MegaBequerels

• Scan mode: Sinogram

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 3 Minutes

#### 4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

• Input trigger signal: None

- LLD (keV): 435
- ULD (keV): 650

## 4.3 Recons

#### 4.3.1 Recon 1

- Series description: PET WB Image check
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

#### 4.3.2 Recon 2

- $\bullet$  Series description: PET FLT True X 2 mm
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 3
- Subsets: 21
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off