TUMOR_HYPOXI_HUND_LM1.MlAdult_PET5

3. december 2012

Indhold

1	Topogram 1				
	1.1	Routine	l		
	1.2	Scan	1		
2	CT	WB	L		
	2.1	Routine	1		
	2.2	Scan	2		
	2.3	Recons			
		2.3.1 Recon 1	2		
3	Pau	se 3	3		
4	PE	C LM	3		
	4.1	Routine	3		
	4.2	Scan	1		
	4.3	Recons	1		
		4.3.1 Recon 1	1		

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 1024 mm
- Tube position: Top

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 1024 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: 80s shar
- Window: Topogram Body

2 CT WB

2.1 Routine

- Eff. mAs: 11
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- \bullet CTDlvol: 0.7414mGy

• Scan time: 3.360 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

2.2 Scan

• Quality ref. mAs: 170

• Eff. mAs: 11

• kV: 120

 \bullet Scan time: 3.360 s

• Rotation time: 0.500 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 1.5000000

• Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

• Series description: AC CT

• Slice: 3.00

• Kernel: B18f very smooth

• Window: Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 2.000

• No. of images: 111

3 Pause

4 PET LM

4.1 Routine

• Isotope: Cu-64

• Pharm.: ATSM

• Inj. Dose: 10 MegaBequerels

• Inj. date (date/month - year): 03/11 - 2009

• Inj. time: 15:30:00

• Scan mode: List mode

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 60 Minutes

4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: List mode

• Sinogram mode: Trues

• Input trigger signal: ECG

• LLD (keV): 435

• ULD (keV): 650

4.3 Recons

4.3.1 Recon 1

• Series description: PET Cardiac

• Recon range (bed): Automatic/All/dunno

• Output image type: NoRecon

• Recon method: Iterative

• Iterations: 4

• Subsets: 8

• Image size: 168

• Zoom: 2

• Filter: Gaussian

• FWHM (mm): 2

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: On

• Match CT slice location: On

• Save intermediate data: Off