

PROJ_NET_FLT_PET_CT_IV_5.MlAdult

27. september 2012

Indhold

1	Topogram	1
1.1	Routine	1
1.2	Scan	1
2	CT WB	1
2.1	Routine	1
2.2	Scan	2
2.3	Recon	2
2.3.1	Recon 1	2
2.3.2	Recon 2	3
2.3.3	Recon 3	4
2.3.4	Recon 4	4
3	Pause	5
4	PET WB	5
4.1	Routine	5
4.2	Scan	5
4.3	Recon	6
4.3.1	Recon 1	6
4.3.2	Recon 2	7

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 0 mm
- Tube position: Top

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 0 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: T80s sharp
- Window: Topogram Body

2 CT WB

2.1 Routine

- Eff. mAs: 225
- kV: 14
- CARE Dose4D: On
- CTDIvol: 15.1475mGy
- Scan time: 1.2000000 s

- Delay: 60.000 s
- Slice: 3.00 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 225 grader

2.2 Scan

- Quality ref. mAs: 225
- Eff. mAs: 225
- kV: 120
- Scan time: 14.840 s
- Rotation time: 0.500 s
- Delay: 60.000 s
- Slice: 0.6 mm
- Pitch: 1.2000000
- Direction: Craniocaudal

2.3 Recon

2.3.1 Recon 1

- Series description: AC CT
- Slice: 3.00
- Kernel: "B19f PET very smooth"
- Window: "Abdomen"
- Extended FoV: On
- FoV: 700
- Center X: 0
- Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 324

2.3.2 Recon 2

- Series description: CT WB 3.0 B30f
- Slice: 3.00
- Kernel: "B30f medium smooth"
- Window: "Abdomen"
- Extended FoV: Off
- FoV: 500
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 324

2.3.3 Recon 3

- Series description: CT WB 3.0 SPO cor
- Slice: 3.00
- Kernel: "B30f medium smooth"
- Window: "Abdomen"
- Extended FoV: Off
- FoV: 971
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: 3D
- Recon Axis: Coronal
- Image order: AP
- Recon increment: 3.000
- No. of images: 167

2.3.4 Recon 4

- Series description: CT WB 3.0 SPO sag
- Slice: 3.00
- Kernel: "B30f medium smooth"
- Window: "Abdomen"
- Extended FoV: Off
- FoV: 971
- Center X: 0
- Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: 3D
- Recon Axis: Sagittal
- Image order: LeftToRight
- Recon increment: 3.000
- No. of images: 324

3 Pause

4 PET WB

4.1 Routine

- Isotope: F-18
- Pharm.: FLT
- Inj. Dose: 1 MegaBequerels
- Inj. date (date/month - year): 26/06 - 2012
- Inj. time: 14:13:00
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 3 PtMinute

4.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram

- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

4.3 Recon

4.3.1 Recon 1

- Series description: PET WB Image check
- Recon range (bed): ups..
- Output image type: Corrected
- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

4.3.2 Recon 2

- Series description: PET FLT Corr True X 2 mm
- Recon range (bed): ups..
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 3
- Subsets: 21
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off