

PETCT_WholeBody_XL.MlAdult_PET5

11. januar 2013

Indhold

1	Topogram	1
1.1	Routine	1
1.2	Scan	1
2	CT WB	1
2.1	Routine	1
2.2	Scan	2
2.3	Recons	2
2.3.1	Recon 1	2
2.3.2	Recon 2	3
2.3.3	Recon 3	4
3	Pause	4
4	PET WB	4
4.1	Routine	4
4.2	Scan	5
4.3	Recons	5
4.3.1	Recon 1	5
4.3.2	Recon 2	6

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 1024 mm
- Tube position: Top

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 1024 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: 80s shar
- Window: Topogram Body

2 CT WB

2.1 Routine

- Eff. mAs: 170
- kV: 120
- CARE Dose4D: On
- CareDoseType: CareDoseAEC
- CTDIvol: 11.4779mGy

- Scan time: 18.670 s
- Delay: 4.000 s
- Slice: 5.00 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

2.2 Scan

- Quality ref. mAs: 170
- Eff. mAs: 170
- kV: 120
- Scan time: 18.670 s
- Rotation time: 0.500 s
- Delay: 4.000 s
- Slice: 5.00 mm
- Pitch: 0.8000000
- Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

- Series description: tmp CT WB RTD
- Slice: 5.00
- Kernel: B19f PET very smooth
- Window: Abdomen
- Extended FoV: On
- FoV: 700
- Center X: 0

- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 11.500
- No. of images: 72

2.3.2 Recon 2

- Series description: AC CT
- Slice: 5.00
- Kernel: B19f PET very smooth
- Window: Abdomen
- Extended FoV: On
- FoV: 700
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 274

2.3.3 Recon 3

- Series description: CT WB 3.0 B30f
- Slice: 3.00
- Kernel: B30f medium smooth
- Window: Abdomen
- Extended FoV: Off
- FoV: 500
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 411

3 Pause

4 PET WB

4.1 Routine

- Isotope: F-18
- Pharm.: FDG
- Inj. Dose: 10 Millicuries
- Inj. date (date/month - year): 30/10 - 2006
- Inj. time: 10:30:00

- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: 5
- Scan duration/bed: 2 Minutes

4.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

4.3 Recons

4.3.1 Recon 1

- Series description: PET WB
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative
- Iterations: 2
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5

- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

4.3.2 Recon 2

- Series description: PET WB Uncorrected
- Recon range (bed): Automatic/All/dunno
- Output image type: Uncorrected
- Recon method: Iterative
- Iterations: 2
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: Off
- Match CT slice location: On
- Save intermediate data: Off