# 8\_PET\_Breathold.Adult

23. oktober 2012

# Indhold

1	Topogram 1			
	1.1	Routine	1	
	1.2	Scan	1	
<b>2</b>	Lunge			
	2.1	Routine	1	
	2.2	Scan	2	
	2.3	Recons	2	
		2.3.1 Recon 1	2	
		2.3.2 Recon 2	3	
		2.3.3 Recon 3	4	
3	Pau	se	4	
4	PE	T Lunge	4	
	4.1	Routine	4	
	4.2	Scan	5	
	4.3	Recons	5	
		4.3.1 Recon 1	5	
		4.3.2 Recon 2	6	
		4.3.3 Recon 3	7	
5	Pau	se	7	
6	PET Lunge			
	6.1	Routine	7	
	6.2	Scan	8	
	6.3	Recons	8	
		6.3.1 Recon 1	8	
		6.3.2 Recon 2	9	
		6.3.3 Recon 3	10	
7	Pau	se	10	
8	PE	T Lunge	10	
	8.1	Routine	10	
	8.2	Scan	11	
	8.3	Recons	11	
		8.3.1 Recon 1	11	
		8.3.2 Recon 2	12	
		8.3.3 Recon 3	13	

# 1 Topogram

# 1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 512 mm

• Tube position: Top

### 1.2 Scan

• mA: 35

• kV: 120

• Delay: 6s

• Topogram length: 512 mm

• Direction: Craniocaudal

• Tube position: Top

• API: API 962494eb-db00-4e54-b01b-9e42ecd67603

• Kernel: 20

• Window: Topogram

# 2 Lunge

# 2.1 Routine

• Eff. mAs: 225

• kV: 120

• CARE Dose4D: On

• CareDoseType: CareDoseAEC

• CTDlvol: 17.1mGy

• Scan time: 0.970 s

• Delay: 10.000 s

• Slice: 3 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

# 2.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 225

• kV: 120

 $\bullet$  Scan time: 0.970 s

• Rotation time: 0.500 s

• Delay: 10.000 s

• Slice: 3 mm

• Pitch: 1.40

• Direction: Caudocranial

# 2.3 Recons

#### 2.3.1 Recon 1

• Series description: tmp Lunge RTD

• Slice: 5

• Kernel: B18f

• Window: Mediastinum

• Extended FoV: false

• FoV: 331

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Caudocranial

• Recon increment: 5.000

• No. of images: 1

#### 2.3.2 Recon 2

• Series description: CT Lunge 3.0 B18f

• Slice: 3

• Kernel: B18f

• Window: Mediastinum

• Extended FoV: false

• FoV: 331

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 1

• Series description: CT Lunge 3.0 B70f

• Slice: 3

• Kernel: B70f

• Window: Lung

• Extended FoV: false

• FoV: 331

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 1

# 3 Pause

# 4 PET Lunge

# 4.1 Routine

• Isotope: F-18

• Pharm.: Fluorodeoxyglucose

• Inj. Dose: 300000000.000000 Bequerels

• Inj. date (date/month - year): 14/05 - 2012

• Inj. time: 14:22:33

- Scan mode: Sinogram
- Scan range: Do not match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 20.000000 Seconds

# 4.2 Scan

- Autoload: true
- Rebinner LUT: N/A
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- ULD (keV): N/A

#### 4.3 Recons

### 4.3.1 Recon 1

- Series description: PET Lunge Iterativ
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 8
- Image size: 200
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4.000000

- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: false
- Match CT slice location: false
- Save intermediate data: false

#### 4.3.2 Recon 2

- Series description: PET Lunge TrueX+TOF
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 3
- Subsets: 24
- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: false
- Match CT slice location: false
- Save intermediate data: false

- Series description: PET Lunge OSEM
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 8
- Image size: 256
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: false
- Match CT slice location: false
- Save intermediate data: false

# 5 Pause

# 6 PET Lunge

# 6.1 Routine

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- Inj. Dose: 300000000.000000 Bequerels
- Inj. date (date/month year): 14/05 2012

• Inj. time: 14:22:33

• Scan mode: Sinogram

• Scan range: Do not match CT FOV

• No. of beds: Not given. Check recon range

• Scan duration/bed: 20.000000 Seconds

# 6.2 Scan

• Autoload: true

• Rebinner LUT: N/A

• Scan output: Sinogram

• Sinogram mode: false

• Input trigger signal: None

• LLD (keV): N/A

• ULD (keV): N/A

# 6.3 Recons

# 6.3.1 Recon 1

• Series description: PET Lunge Iterativ

• Recon range (bed): Automatic/All/dunno

• Output image type: Corrected

• Recon method: Iterative3D

• Iterations: 4

• Subsets: 8

• Image size: 200

• Zoom: 1

• Filter: Gaussian

- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: false
- Match CT slice location: false
- Save intermediate data: false

#### 6.3.2 Recon 2

- Series description: PET Lunge TrueX+TOF
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 3
- Subsets: 24
- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: false
- Match CT slice location: false
- Save intermediate data: false

- Series description: PET Lunge OSEM
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 8
- Image size: 256
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: false
- Match CT slice location: false
- Save intermediate data: false

# 7 Pause

# 8 PET Lunge

# 8.1 Routine

- Isotope: F-18
- $\bullet$  Pharm.: Fluorodeoxyglucose
- Inj. Dose: 300000000.000000 Bequerels
- Inj. date (date/month year): 14/05 2012

• Inj. time: 14:22:33

• Scan mode: Sinogram

• Scan range: Do not match CT FOV

• No. of beds: Not given. Check recon range

• Scan duration/bed: 20.000000 Seconds

# 8.2 Scan

• Autoload: true

• Rebinner LUT: N/A

• Scan output: Sinogram

• Sinogram mode: false

• Input trigger signal: None

• LLD (keV): N/A

• ULD (keV): N/A

# 8.3 Recons

# 8.3.1 Recon 1

• Series description: PET Lunge Iterativ

• Recon range (bed): Automatic/All/dunno

• Output image type: Corrected

• Recon method: Iterative3D

• Iterations: 4

• Subsets: 8

• Image size: 200

• Zoom: 1

• Filter: Gaussian

- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: false
- Match CT slice location: false
- Save intermediate data: false

#### 8.3.2 Recon 2

- Series description: PET Lunge TrueX+TOF
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 3
- Subsets: 24
- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: false
- Match CT slice location: false
- Save intermediate data: false

- Series description: PET Lunge OSEM
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 8
- Image size: 256
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: false
- Match CT slice location: false
- Save intermediate data: false