$PETCT_NEMA_2007_IQ_MOD.Adult$

23. oktober 2012

Indhold

| 1 | Topogram | | | | |
|----------|---------------|---------------|---|--|--|
| | 1.1 | Routine | 1 | | |
| | 1.2 | Scan | 1 | | |
| 2 | \mathbf{AC} | CT NEMA | 1 | | |
| | 2.1 | Routine | 1 | | |
| | 2.2 | Scan | 2 | | |
| | 2.3 | Recons | 2 | | |
| | | 2.3.1 Recon 1 | 2 | | |
| | | 2.3.2 Recon 2 | 3 | | |
| 3 | Pau | se | 4 | | |
| 4 | PE | NEMA IQ | 4 | | |
| | 4.1 | Routine | 4 | | |
| | 4.2 | Scan | 4 | | |
| | 4.3 | Recons | 5 | | |
| | | 4.3.1 Recon 1 | 5 | | |

1 Topogram

1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 1024 mm

• Tube position: Top

1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 1024 mm

• Direction: Craniocaudal

• Tube position: Top

• API: API

• Kernel: 80

• Window: Topogram

2 AC CT NEMA

2.1 Routine

• Eff. mAs: 50

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

• CTDlvol: 3.64mGy

• Scan time: 8.050 s

• Delay: 2.000 s

• Slice: 5 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

2.2 Scan

• Quality ref. mAs: 120

• Eff. mAs: 50

• kV: 120

 \bullet Scan time: 8.050 s

• Rotation time: 0.500 s

• Delay: 2.000 s

• Slice: 5 mm

• Pitch: 0.80

• Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

• Series description: tmp AC CT NEMA RTD

• Slice: 4

• Kernel: B19f

• Window: Abdomen

• Extended FoV: false

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 4.000

• No. of images: 56

2.3.2 Recon 2

• Series description: AC CT NEMA

• Slice: 5

• Kernel: B19f

• Window: Abdomen

• Extended FoV: true

• FoV: 780

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 75

3 Pause

4 PET NEMA IQ

4.1 Routine

• Isotope: F-18

• Pharm.: FDG

• Inj. Dose: 370000000.000000 Bequerels

• Inj. date (date/month - year): 09/09 - 2010

• Inj. time: 11:01:07

• Scan mode: Sinogram

• Scan range: Do not match CT FOV

• No. of beds: 0

• Scan duration/bed: 540.000000 Seconds

4.2 Scan

• Autoload: true

• Rebinner LUT: N/A

• Scan output: Sinogram

• Sinogram mode: PromptsRandoms

• Input trigger signal: None

• LLD (keV): N/A

• ULD (keV): N/A

4.3 Recons

4.3.1 Recon 1

- Series description: PET NEMA IQ
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 3
- Subsets: 24
- Image size: 200
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: true
- Match CT slice location: false
- Save intermediate data: true